



# POLISH CENTRE FOR TESTING AND CERTIFICATION

02-699 Warszawa, ul. Kłobucka 23A

**Mechanical Laboratory**

tel.: (+48 22) 46 45 594; fax: (+48 22) 46 45 563

e-mail: labmech@pcbc.gov.pl

FRL-15  
03/09

Date: 15.12.2010

## TEST REPORT NO. BR – 200/L-233/2010

Page 1 of 4

<b>Subject of testing:</b>	<i>Shower stool</i>	<b>Classification according to PN-EN ISO 9999:2007 :</b>	<b>09 33.03</b>
<b>Type/Model</b>	<b>302064</b>	<b>Catalogue no.:</b>	<b>2141</b>
<b>Manufacturer :</b>	<b>MOBILEX A/S, Nørskovvej 1 DK - 8660 Skanderborg</b>	<b>Number of specimens :</b>	<b>1</b>
<b>Applicant:</b>	<b>A-Net s.c. 93-469 Łódź, ul. Łaskowice174</b>		
<b>Kind of testing</b>	<b>Testing scope according to application of Client Mechanical testing for conformity with PN-EN 12182:2005 excluding clause 4.1, 4.3, 5.2, 22, 24</b>		

**Test started:** 02.12.2010

**Test finished :** 15.12.2010.

**Performed by:**

**Mirosław Szymański**

**Checked by :**

**Ireneusz Czerwiński**

**Approved by :**

KIEROWNIK  
LABORATORIUM MECHANICZNEGO

**mgr inż. Andrzej Tkaczyk**

**Special comments / enclosures:**

Copyright © 2005 by Polish Centre for Testing and Certification (applicable to report form)

Test results refer only to tested units.

Test results reported here are not applicable to the further modifications of the product affecting its structure, material or technology.

This test report shall be neither copied differently as in the whole nor be published without written consent of the Laboratory.



## TESTING

## NORMATIVE REFERENCES

PN-EN 12182:2005

Technical aids for disabled persons – General requirements and test methods

## PHOTO OF PRODUCT



## CHARAKTERISTICS OF PRODUCT

<b>Name of product:</b> <i>shower stool</i>	<b>Maximum weight of user:</b> <i>130kg</i>	<b>Mass of product:</b> <i>1,5 kg</i>
<b>Dimensions:</b>		
<b>Owerall high</b>	<i>355÷535 mm</i>	
<b>Diameter of seat</b>	<i>330 mm</i>	
<b>Spacing of stand</b>	<i>480 x 500 mm</i>	
<b>Material:</b>		
<i>Aluminium alloy, plastic</i>		