

CENTRE FOR TESTING AND CERTIFICATION - MECH-TEST

Mechanical Laboratory

05-077 Warszawa-Wesoła, ul. Klonowa 22 tel.: +48 603 23-26-45, e-mail: cbc.mech.test@gmail.com, www.cbc.org.pl

Date 20.08.2018

TEST REPORT NO. *CBC* -075/2018

Page 1 of 4

Subject of testing:

Walking aids with built-in handgrips and three or

more legs of which two or more are having wheels.

which provide support whilst walking

Type / Model:

Balance rollator

Nr kat .: --

Art. Nr.: 312481

312482 312483 312484 312485

Number of specimens: 1

Classification according to PN-EN ISO 9999:2011:12 06 06

Manufacturer:

MOBILEX A/S

Grønlandsvej 5

DK - 8660 Skanderborg

Applicant:

A-Net s.c.

93-469 Łódź. ul. Łaskowice174

Kind of testing

Mechanical testing for conformity with PN-EN ISO 11199-2: 2005

Test started: 31.07.2018

Test finished: 20.08.2018

Approved by:

inż. Andrzej Tkaczyk

Special comments / enclosures:

Copyright © 2012 by Centre for Testing and Certification (applicable to report form)

Test results refer only to tested units.

Test results reported here are not applicable to the further modifications of the product affecting its structure, material or technology. This test report shall be neither copied differently as in the whole nor be published without written consent of the Laboratory.



CHARACTERISTIC OF PRODUCT

Name: Balance rollator **Dimension of rollator:**

Art. nr.: 312481, 312482, 312483, 312484, 312485 SN: --

Maximum permissible user mass: 120 kg

Mass of rollator: Description

DITAMA	OF	DDA	TAT	TATE
PHOTO	E DH.	P 12 ()		14
THOTO	OI.	TILO	\mathbf{v}	

8,08 kg

	Descripti	on	
Ele	ments/parameters/materials/di	mensions	Comments
	Distance between	438/460	
15	handgrips (dimension 2)	mm	
late (2	Angle between of handgrip	$ 4^0 $	
rol 99-	axis and direction of		
ng 11	movement (α)		
Iki O	Height of rollator	744 mm	min.
wa	(dimension 6)	885 mm	max.
P di	Width of rollator	627 mm	
ns PN	(dimension 5)		
sio	Turning width	632 mm	
Dimensions od walking rollator (fig. 2 PN-EN ISO 11199-2)	(dimension 1)		
) in	Length of rollator	628 mm	
_	(dimension 4)		
Dimen	sions of folded rollator (mm)	807 x 627 x	260
	Handgrip - diameter	27x41 mm	Anatomical
60		32x45 mm	handgrip
Fig.	Handgrip - length	120/80 mm	
	Front wheels- quantity	2	castor
atol	Front wheels - diameter	161 mm	wheels
ollo	Front wheels – width	26/29 mm	
f r	Front wheels - brake	none	
S O	Rear wheels - quantity	2	
eel	Rear wheels - diameter	161 mm	
Wheels of rollator	Rear wheels - width	26/29 mm	
	Rear wheels - brake	Included	
Tip	Diameter		
	Material	Not any	
	Colour		
	Front legs	Aluminum,	
ılla	Bracing member (no. 8)	Steel,	
f rc	Rear legs	Hard plastic	2,
10	Height adjusting device (no.	Bolts, nuts	
eria fig.	4)		
Material of rollator (fig. 1)	Handgrip (no 5),	Hard plastic	
5 5	Brake elements	•	





MOBILEX A/S Grønlandsvej 5 DK - 8660 Skanderborg Tel: +45 87 93 22 20

Balance rollator For indoor use only

item nr. 312481 (red) - 312482 (grey)

Max. 120 kg

Max. width = 63 cm





Produced

www.mobilex.dk

Serial no.









Report no.: CBC-075/2018 Page: 3 of 4

Requirements according to clause	Test me- thod according to clause		Checked characteristics/assemblies/parameters		Real	Test result	Comments	
4.1 Measur. Manoeuvrability				0 101 m		ø 161 mm width 26mm Conf.	Pos.	ø front wheels ≥75mm outdoor intended rollato ø front wheels ≥180mm width of wheels ≥28mm
4.2	5.3			ection stability		18,0 ° Conf.	Pos.	Stability required ≥ 1
4.2	5.4	Backward-direction stability				12,8 ° Conf.	Pos.	Stability required ≥
4.2	5.5			rection stability	To 1	4,4 ° Conf.	Pos.	Stability required ≥ 3,
4.2	5.6		ility –	hasket has (5kg) drin avvsen	forwards		N/A N/A	Stability required ≥ 1.
		with loaded basket, bag (5kg), drip, oxygen cylinder backwards					N/A N/A	Stability required ≥ 3. Stability required ≥ 3.
4.3	V/I			ng facility during rollator motion		Conf.	Pos.	Sidotaly required 23.
A common or a management	V/I		resting	g brakes in rollator with more than seat or intended for outdoor use	n 2 wheels and	Conf.	Pos.	
[5.7.1.1			grip distance (fig. 4, dimension 1)			N/A	≤ 75 mm
	5.7.1		Runnin	g brake effectiveness		Conf.	Pos.	Movement of rollator ≤ 10 mm in 1 minute
	Measur.	Brakes	Force to	o set parking brake			D	
-	Measur.	Bra		15		40 N Conf.	Pos.	≤ 60 N
1				o release parking brake		30 N Conf.	Pos.	≤ 40 N
5.7.2				g brake effectiveness		Conf.	Pos.	Movement of rollator ≤ 10 mm in 1 minute
	V/I			lity to compensate brake wear		Conf.	Pos.	
	V/I		adjustir	not adversely affected by folding, ing actions of rollator	unforlding or	Conf.	Pos.	
4.4	Measur. V/I	Handgrip				27 mm/ 32mm Conf.	Pos.	Width of handgrip ≥20i and ≤50mm
4.5	Measur. V/I	Leg section and tip					N/A	ø tip ≥35mm (tested roll is equipped in four whe
4.6	5.10	Rest	ing seat	– static loading durability	Conf.	Pos.	I minute under load 1,2 x user's weight±29 (144kg)	
4.7	5.12		echanica urability			Conf.	Pos.	200 000 cycles with loc 96kg±2%, f=1Hz
4.7	5.11			Static loading test		Conf.	Pos.	loading 144kg±2%, 5sek.
4.8	V/I 5.14		isting de			Conf.	Pos.	
4.11	ISO 10993-	rold		namsin Biocompatibility of material with	human hody	Conf.	Pos.	
7.11	1	Sisteman	116	Diocompationity of material with	numan body		N/T	
	V/I V/I V/I			Free of discolouring of skin or clowith rollator materials	othing in contact	Conf.	Pos.	
V/I		2 0	3	Burrs, shar edges, projections		Conf.	Pos.	
2	X 7 / Y				pelling of product	T 2 2 2 2 1		
5.2	V/I			user mass		Included	Pos.	
	}			safe working load (SWL) to be r			N/A	
	 c) Maximum allowed angle between the longitudinal centreline of handle and the direction of motion, if the handles are sideways adjustable 				N/A			
		d) Manufacturer's name or trade name and address			Included	Pos.		
		e) Manufacturer's model identification name and/or number f) Month and year of manufacture				Included	Pos.	
						Included	Pos.	
		g) Maximum extension of the height adjustment, marked on the adjusting members h) Maximum width of the rollator			Included	Pos.		
	-					Included	Pos.	
	Ì			tended for outdoor/indoor use		Included		
.10	V/I	Warning showing allowed angle between handle axis and direction movement or physical stop of angle adjusting				Pos. N/A		



Mechanical Laboratory of CBC				Report no.: CBC-075/2018			
					Page: 4 of 4		
		Contents of user manual and/or assembly manual or clear and indeli	ble markin	g of produ	ct		
6.3	V/I	a) Maximum rollator height		N/T			
ŀ		b) Minimum rollator height		N/T			
		c) maintenance and cleaning instructions, including a description of the method and suitable cleaning agents and any precautions needed to avoid corrosion and/or ageing of the materials used in construction of the rollator		N/T			
		d) Instructions for assembly, adjustment of all kinds, folding and unfolding		N/T			
		e) Warnings and advice about precautions relating to safe distances between moving and stationary parts (see EN 12182, Clauses 12 and 13, for guidance)		N/T			
		 f) Maximum safe working load (SWL) for load carrying accessories such as basket, tray, shopping bag, etc. 		N/T			
4.10	V/I	Warning in user manual on consequences of such an adjustment of angle between handle longitudinal axis and direction of movement outside allowed value (when handles are adjustable aside).		N/A			
		TEST CONDITIONS					
Ambient temperature		19°C		Required temperature 21°C ±5°C			
Relative humidity of air:		55 %		Not required			
Commer	nts:						
		n maximum height adjustment of rollator.					
Il tests j	performend in	the least stabble position of self-adjusting wheels.			25,000		
ests per	formed with ha	andles positioned at their maximum (allowed) angle to the direction of moti	on (when ac	djustment is	s possible).		
		lity test, static loading test, fatigue test.					
	tor was tested.						
Juring v	isual inspect	ion before testing any visible defects that could have influence on te	st results v	were not st	tated.		

Pos. – positive; Neg – negative; N/T – not tested; N/A – not applicable; N/R – not required , N/O – not occurred , V/I. – visual inspection, Conf. – conformed.

CONCLUSIONS:

Testing object conforms with requirements of PN-EN ISO 11199-2: 2005 in scope of mechanical testing ordered by client, excluding biocompatibility tests of material with human body according to PN-EN ISO 10993-1:2010

---- END ----

