



POLISH CENTRE FOR TESTING AND CERTIFICATION

02-699 Warszawa, ul. Kłobucka 23A

Mechanical Laboratory

tel.: (+48 22) 46 45 594; fax: (+48 22) 46 45 563

e-mail: labmech@pcbc.gov.pl

Date: 12.04.2010

TEST REPORT NO. BR – 055/L-071/2010

Page 1 of 4

Subject of testing: *Shower Chair with tilt*

Classification according to PN-EN ISO 9999:2001 :
09 33 03

Type/Model: 302024

Catalogue no.: ---

Manufacturer : *MOBILEX A/S, Noerskovvej 1
DK - 8660 Skanderborg*

Number of specimens : 1

Applicant: *A-Net s.c. 93-469 Łódź,
ul. Łaskowice174*

Kind of testing: *Testing scope according to application of Client
Mechanical testing for conformity with PN-EN 12182:2005
excluding clause 4.1, 4.3, 5.2, 22, 24*

Test started: 6.04.2010

Test finished : 12.04.2010

Performed by:

Checked by :

Approved by :

Mirosław Szymański

Ireneusz Czerwiński

KIEROWNIK
LABORATORIUM MECHANICZNEGO

mgr inż. Andrzej Tkaczyk

Special comments / enclosures:

Copyright © 2005 by Polish Centre for Testing and Certification (applicable to report form)

Test results refer only to tested units.

Test results reported here are not applicable to the further modifications of the product affecting its structure, material or technology.

This test report shall be neither copied differently as in the whole nor be published without written consent of the Laboratory.





TESTING

NORMATIVE REFERENCES

PN-EN 12182:2005

Technical aids for disabled persons – General requirements and test methods

PHOTO OF PRODUCT



MOBILEX A/S
Noerskovvej 1
DK - 8660 Skanderborg
Tel: +45 87 93 22 20
www.mobilex-care.com

CE Produced Serial no.

Shower chair
with tilt



Max. 130 kg



Warning!
Be Careful!
Risk of finger cut!



CHARAKTERISTICS OF PRODUCT

Name of product: Shower Chair with tilt **Maximum weight of user:** 130 kg **Mass of product:** 15,1 kg

Dimensions:

Overall height with headrest:	970 mm – 1410 mm
Overall height without headrest:	850 mm – 1110 mm
Maximum width:	558 mm
Maximum length (max. moving of footrest and headrest):	1310 mm
Minimum length(min. moving of footrest and headrest):	860 mm
Height of seat:	585 mm/535 mm (2 levels of adjustment)
Angle of seat plane:	from - 6,5 ⁰ to + 25,5 ⁰

Material:

Aluminum alloy, plastic