

Delichon Ltd, 1 The Clayfields, Sandleheath Industrial Estate, Old Brickyard Road, Fordingbridge, Hampshire SP6 1PA **T:** +44 (0) 1725 519405 **E:** office@delichon.uk www.delichon.uk

Test Standards for Delta Buggy product range manufactured by Delichon Ltd.

Delta All-Terrain Buggy sizes: S, M1, M2, L1, L2, XL Delta Recline All-Terrain Buggy sizes: S, M, L Delta Run All-Terrain Buggy sizes: M1, M2, L1, L2, XL

The above models have been independently tested and approved according to the following standards:

Test Standard	Date of Test	Testing Centre
EN 12182:2012	Nov 2018	TUV Rhineland
EN 12183-7.2:2014	Nov 2018	TUV Rhineland
EN 12183-9.2.2.3:2014	Nov 2018	TUV Rhineland
EN 1021-1:2014	April 2018	BLC Leather Technology Centre, UK
EN ISO 10993-1:2009	Jan 2017	HYG Centre, Germany
EN ISO 10993-5:2009	Jan 2017	HYG Centre, Germany
ISO 7176-8:2014	Jan 2016	TUV Rhineland

Registered in England & Wales No. 3184535 | VAT Reg. No. GB 675 7325 04 Registered Office:1 The Clayfields, Sandleheath Industrial Estate, Old Brickyard Road, Fordingbridge, Hampshire SP6 1PA



Delichon Delta

Section 1-Classification and Declaration

1.1 Classification

The Delta Buggy is a non-invasive device in accordance with ANNEX I X of the Medical Devices Directive 93/42EEC (MOD), Classification Rule 1 is therefore subject to self-certification as a Class 1 device.

1.2 Declaration of Conformity

As the manufacturer and distributor of the

Delta Buggy

Delichon Ltd 1 The Clayfields, Sandleheath Industrial Estate, Fordingbridge, Hampshire SP6 1PA Tel: +44 (0)1725 519405

Declares that the product known as the Delta Buggy is a class 1 product in compliance with:

The UK Medical Devices Regulations 2002 Statutory Instrument 2002 No. 618

& The Medical Devices Directive 93/42EEC of 14th June 1993

Signed:

Martin Davy ManagingDirector Delichon Ltd

Date: 01-02-2020

Registered in England & Wales No. 3184535 | VAT Reg. No. GB 675 7325 04 Registered Office:1 The Clayfields, Sandleheath Industrial Estate, Old Brickyard Road, Fordingbridge, Hampshire SP6 1PA