



Product Service

CERTIFICATE

No. Z1 084376 0003 Rev. 00

Holder of Certificate: **Pressalit A/S**

Pressalitvej 1
8680 Ry
DENMARK

Certification Mark:



Product:

**Sanitary aids
PLUS Support arm system**

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition, the certification holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. All applicable requirements of the testing and certification regulations of TÜV SÜD Group have to be complied. For details see: www.tuvsud.com/ps-cert

Test report no.: 713183470

Valid until: 2029-01-28

Date, 2024-02-07

(Torsten Zimmer)

CERTIFICATE

No. Z1 084376 0003 Rev. 00

Model(s): R360070/R360085, R381470/R381485, R370370/R370470, R370385/R370485, R371370/R371470, R371385/R371485, R391370/R391385, R391470/R391485

Brand Name: Pressalit

Parameters:

Type	Handle diameter	Max. length	Mounting plate	Max. user weight
R360070	50 mm	700 mm	210 x 100 mm	150 kg
R360085		850 mm		
R370370		730 mm	150 (424) x 128 mm	
R370470		730 mm		
R370385		880 mm		
R370485		880 mm		
R371370		730 mm		
R371470		730 mm		
R371385		880 mm		
R371485		880 mm		
R391370		770 mm		
R391470		770 mm		
R391385		920 mm		
R391485		920 mm		
R381470		730 mm		
R381485		880 mm		

Tested according to: ISO 17966:2016