

Test Report

Report No.: 923154



**DANISH
TECHNOLOGICAL
INSTITUTE**

Gregersensvej
DK-2630 Taastrup
Tel. +45 72 20 20 00
Fax +45 72 20 20 19

info@teknologisk.dk
www.teknologisk.dk

Assignor: Winnicare Nordic Aps,
Taarnborgvej 12C
DK-4220 Korsør

Page 1 of 1
jlj/moso/hbs
Order no. 923154
No. of appendices: 2

Subject: Model: Triton/Dual Shower commode chair used for showering and toilet.

Type:	Assistive products for personal hygiene that support users. Expected lifetime: 10 years				
Item:	044-00051		Main (lift) actuator	See photos in appendix 2.	
SN:	TRT-200046		Tilt actuator		
Weight:	41 kg	SWL:	150 kg	Production date:	30-03-2020
Materials:	Painted steel, aluminium, plastic and electronics, wheel diameter 100 mm				

Sampling: The test material was sampled by the client and received at the Danish Technological Institute 08-04-2020.

Method: ISO 17966:2016 Assistive products for personal hygiene that support users – Requirements and test methods

Period: The testing was carried out from 08-04-2020 to 25-06-2020.

Result: Model Triton/Dual Commode shower chair meet the requirements of all the tested parts of ISO 17966:2016.

Individual results appear from Appendix 1.

Storage: The test material will be destroyed after 1 month, unless otherwise agreed.

Terms: Accredited testing was carried out in compliance with international requirements (EN/ISO/IEC 17025:2005) and in compliance with Danish Technological Institute's (DTI) General Terms and Conditions regarding Commissioned Work accepted by Danish Technological Institute. The test results apply to the tested products only. This report may be quoted in extract only if the laboratory has granted its written consent.

Date/place: 25-06-2020, Danish Technological Institute, Wood and Biomaterials, Taastrup

Signature: Test responsible

Co-signatory

Digitally signed by:
Jens Ljørring
Ph.-Direct: +45 72202326
E-mail: jlj@teknologisk.dk

Morten
Skotte
Sørensen

Digitalt signeret af
Morten Skotte Sørensen
Dato: 2020.06.26
08:49:50 +02'00'

