

## Manufacturer's declaration of conformity

Care of Sweden AB Name and address

Visit: Fabriksgatan 5A, SE-514 33 Tranemo,

Mail: Box 146, SE-514 23 Tranemo

Phone: +46 (0)771 106 600, Fax +46 (0)325 128 40

info@careofsweden.se

SE-MF-000014315 SRN

**Product group** Curera Seat cushions

7331345PC050CE160HF **Basic UDI-DI** 

**EMDN** code V08030102

Classification Class I, according to Annex VIII (MDR (EU) 2017/745, rule 1

Product(s)/Device name Curera Contur Child

Curera Contur

Curera Contur EX

Curera Hemi right

Curera Hemi left

Curera Hemi double

Curera Plantur

Curera Consept

Curera Consept EX

**Curera Consept Latus** 

Curera Consil

CuroCell S.A.C.

**Curera Seat Cushion** 

Curera Tremix

Configuration The Seat cushions are CE-marked together with seat cushion covers.

Mark of compliance

**Declaration** 

We declare under our sole responsibility as Manufacturer that the product(s) listed above conform to the requirements of the MDR (EU) 2017/745. The product(s) meet(s) the relevant General Safety and

Performance Requirements of Annex I.

The conformity assessment procedure was performed following Annex II to III of MDR (EU) 2017/745. Any modification to the device,

not authorized by us, will invalidate this declaration.

Valid from: 2021-05-01

**Updated:** 2023-02-03

Manufacturer's signature

Susanne Andersson, Chief QA and Compliance Officer

**Date** 2023-02-03

Version: 3.0 1(1)