



Human Handling Concept



### EU DECLARATION OF CONFORMITY

Name and address of the firm: master care A/S  
Sofienlystvej 3, 8340 Malling, Denmark

Single Registration Number (SRN): DK-MF-000011791

We, as the manufacturer, declare that we have issued this EU declaration of conformity under our sole responsibility and that the products listed below, all class 1 according to annex VIII of MDR 2017/745 – rule 1, meets all the provisions of the MDR 2017/745 which apply to conformity assessment procedure MDR Art. 52 & Annex II, III (Class I).

Classification: Medical Device, Classification 1.

Product name: Master Turner

Basic-UDI-DI: 57440019763541QW

Article no.:

MC 001-1303, MC 001-1305, MC 001-1500, MC 001-1700, MC 001-1701, MC 001-1705, MC 001-1706, MC 001-1710, MC 001-1711, MC 001-1712, MC 001-1713, MC 001-1723, MC 001-1725, MC 001-1732, MC 001-1734, MC 001-1735, MC 001-1736, MC 001-1737, MC 001-1738, MC 001-1739, MC 001-1740, MC 001-1747, MC 001-1750, MC 001-1760, MC 001-1770, MC 001-1775, MC 001-1780, MC 001-1912, MC 001-1922, MC 001-1926, MC 001-2000, MC 001-2001, MC 001-2040, MC 001-2500, MC 001-2501, MC 001-2502, MC 001-2503, MC 001-2517, MC 001-2525, MC 001-2535, MC 001-2536, MC 001-2537, MC 001-2545.

Intended use:

The Master Turner and the Cover form a turning system that can be used for users where moving around in bed is difficult either due to physical or cognitive deficits. Friction is reduced between the user and the surface, thereby reducing the risk of shear injury for user and optimizing the ergonomic conditions for the caregiver when turning, moving the user up/down in bed or transferring to sitting on the bed side. The Master Turner should always be used in combination with a Mattress Cover from master care A/S to insure the best reduction of friction.

Declaration's validity:

5 years after signing or changes to the technical documentation.

Date of Issue and place: Malling, 15.01.2025

Name and function:

Vidar Geitheim, CEO