care of sweden

Manufacturer's declaration of conformity

| Name and address | Care of Sweden AB Visit: Fabriksgatan 5A, SE-514 33 Tranemo, |
|--------------------------|--|
| | Mail: Box 146, SE-514 23 Tranemo |
| | Phone: +46 (0)771 106 600, Fax +46 (0)325 128 40 |
| | info@careofsweden.se |
| SRN | SE-MF-000014315 |
| Product group | Optimal preventative mattresses |
| Basic UDI-DI | 7331345PC030CE130G4 |
| EMDN code | V080702 |
| Classification | Class I, according to Annex VIII (MDR (EU) 2017/745, rule 1 |
| Product(s)/Device name | Optimal 5zon Optimal 5zon BM Optimal 5zon Plus Optimal 5zon Eco Apollo 6 Optimal Care Optimal Cura Optimal Cura 12 Optimal Mono Optimal Solett Optimal Solett BM Optimal Maxx Prima 12 |
| Configuration | The mattresses are CE-marked together with a mattress cover. |
| Mark of compliance | CE |
| Declaration | We declare under our sole responsibility as Manufacturer that the product(s) listed above conform to the requirements of the MDR (EU) 2017/745. The product(s) meet(s) the relevant General Safety and Performance Requirements of Annex I. The conformity assessment procedure was performed following Annex II to III of MDR (EU) 2017/745. Any modification to the device, not authorized by us, will invalidate this declaration. |
| Valid from: | 2021-05-01 |
| Updated: | 2023-02-03 |
| Manufacturer's signature | Frankfrhu |
| | Susanne Andersson, Chief QA and Compliance Officer |
| Date | 2023-02-03 |