



Human Handling Concept



EU DECLARATION OF CONFORMITY

Name and address of the firm: master care A/S
Sofienlystvej 3, 8340 Malling, Denmark
www.master-care.dk

Single Registration Number (SRN): DK-MF-000011791

We, as the manufacturer, declare that we have issued this EU declaration of conformity under our sole responsibility and that the products listed below, all class 1 according to annex VIII of MDR 2017/745 – rule 1, meets all the provisions of the MDR 2017/745 which apply to conformity assessment procedure MDR Art. 52 & Annex II, III (Class I).

Classification: Medical device, Class I, rule 1 (Annex VIII)
Product name: Madrascover Vario Black
Basic-UDI-DI: 57440019763286R5
Device description: Low-friction mattress cover for repositioning
EMDN code: Z12019010 – Patient transfer systems

Article no. / UDI-DI (GTIN-14):

MC 001-2810 – 05744001975602
MC 001-2812 – 05744001975619
MC 001-2813 – 05744001975626
MC 001-2820 – 05744001975633
MC 001-2822 – 05744001975640
MC 001-2823 – 05744001975657
MC 001-2840 – 05744001975664
MC 001-2841 – 05744001975671
MC 001-2842 – 05744001975688

Intended use:

The intended use for the product is to reduce friction between the user and the mattress to make it easier for the helper and gentler for the user whilst turning, moving up/ downwards, and moving to sitting position on the edge of the bed. This product should always be used in combination with a Master Turner from master care A/S.

Harmonised / applied standards: Separate list available upon request

Declaration's validity:

Valid until superseded by a newer version or until changes to the device or the technical documentation.

Place of issue: Malling, Denmark

Date of issue: 9.9.2026

Name and function:

Vidar Geitheim, CEO


HUMAN HANDLING SYSTEM
master care A/S • Sofienlystvej 3 • DK-8340 Malling
Tlf. +45 93 85 80 • info@master-care.dk
www.master-care.dk

Signature, on behalf of master care A/S