CE

Declaration of Conformity

Manufacturer name:	Danrehab A/S
Manufacturer address:	Vejlevej 15, Ølholm DK-7160 Tørring
Manufacturer SRN:	DK- 66261417
Device Name:	ERGOtip 6 Anytoilet (L)
Device Description:	Toilet-batch chair with seat tilt, electrical height adjustment



Part Number:	148105
Basic UDI:	571506409120323EL:
HMI No.:	148105
Classification:	Class I
Conforms to regulation:	Medical Device Regulation (EU) MDR 2017/745
The Product has been tested, and evaluated in accordance with the following standards:	EN 17966 EN 60601-1:2013 EN 60601-1-2:2015
The product has been designed and manufactured under a certified quality management system in accordance with:	EN ISO 13485:2016

We hereby declare that the medical device defined above complies with the Regulation (EU) MDR 2017/745 for medical devices.

Maria W Laursen

Head of Quality and Compliance, MDR
25-05-2026