Declaration of Conformity

CE

ViewPlus Technologies, Inc.
1965 SW Airport Avenue
Corvallis Oregon 97333 U.S.A.
phone: 541.754.4002 fax: 541.738.6505

We, ViewPlus Technologies, Inc.
1965 SW Airport Ave.
Corvallis, Oregon 97333, USA
declare under our sole responsibility that our products listed below meet these directives and standards:

Authorized European Representative
ViewPlus GmbH
Kolpingweg 15
49492 Westerkappeln, Germany
49 (0) 5404 – 9596 414
e-mail: info@viewplus.com

Application of Council Directive(s): EMC directive 89/336/EEC
Amending Directives: 92/31/EEC,93/68/EEC, 98/13/EEC
Medical Device Regulation: (EU) 2017/745
Low Voltage Directive 2006/95/EC

Test Description	Specification	
Emissions	EN 55032:2012 Class B	CISPR 32:2015
Power Line Flicker	EN 61000-3-3:2013	IEC 61000-3-3:2013
Power Line Harmonics	EN 61000-3-2:2014	IEC 61000-3-2:2014
Electrostatic discharge	EN 55024:2010	IEC 61000-4-2:2008
Radiated immunity	EN 55024:2010	IEC 61000-4-3:2010
Electrical fast transient/burst	EN 55024:2010	IEC 61000-4-4:2012
Surge	EN 55024:2010	IEC 61000-4-5:2014
Conducted immunity	EN 55024:2010	IEC 61000-4-6:2013
Magnetic Field Immunity	EN 55024:2010	IEC 61000-4-8:2009
Input voltage interruptions/DIPS	EN 55024:2010	IEC 61000-4-11:2004
EN12182:2012	Technical aids for disabled persons	IEC 60950-1

During the conformity assessment procedure of the Medical Device Directive, a risk analysis in accordance to EN ISO14971 has been performed for the products listed below.
Class 1 according to annex VII of MDR (EU) 2017/745
Type of Equipment: Information Technology Equipment for Office/Home Office Use
Class B
Model No. 121201 Rogue Trac
Year of Manufacture: from 2021
Model No. 121202 Rogue Sheet
Year of Manufacture: from 2021

I, the undersigned, hereby declare that the equipment specified above conforms to the above Directive(s) and Standard(s).

Place: Corvallis, Oregon USA

Carolyn K Gardner

Carolyn K Gardner
Corporate Secretary

Date: 3 August 2022