



Product Service

# CERTIFICATE

No. Z1 17 08 86271 032

**Holder of Certificate:** ArjoHuntleigh ABHans Michelsensgatan 10  
211 20 Malmö  
SWEDEN**Production Facility(ies):** 65428**Certification Mark:****Product:** Sanitary aids  
Height adjustable shower and toilet chair**Model(s):** Carendo**Parameters:**

Rated voltage	24 Vdc
Rated power	150 VA
Degree of protection	IPX4
Type of applied part	B
Operating time	2 min. / 18 min.
Max. patient weight	136 kg / 300 lbs
Max. dimension	1510 x 640 x 1550 mm

**Tested according to:** IEC 60601-1:2005  
IEC 60601-1:2005/AMD1:2012  
IEC 60601-1-6:2010  
IEC 60601-1-6:2010/AMD1:2013  
IEC 60601-1-11:2015

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

**Test report no.:** 713102336**Valid until:** 2022-08-09**Date,** 2017-08-10  (Torsten Zimmer)

Page 1 of 1

