

# FAABORG REHAB TECHNIC ApS.



Faaborg d. 01-06-2020

## Declaration of CE conformity.

The following hoists are CE - marked complying with the.:

**Announcement regarding Medical Equipment No. 225 of March the 11th 2011**

**DS/EN ISO 10535:2007 Hoists for the transfer of disabled persons – Requirements and test methods and The European Communities Council Directive 93/42/EEC of 14<sup>th</sup> July 1993**

Careline Jutlandia PL -135-1-2  
Careline Jutlandia PL -165-1-2  
Careline Jutlandia PL –Stand up 165-1-2  
Careline Selandia VL – 200-1-2- 2,5  
Careline Selandia VL – 250 – 1-2 / VL-300-1-2  
Careline Selandia VL – Stand up-165-1-2  
Careline King VL- 200/250-1-2-3  
Careline Count VL-200/250-1-2  
Careline Fionia Nordic 95-1-2  
Careline Fionia Nordic 165-1-2  
Careline Fionia Nordic Basic165-1-2-3  
Careline Fionia Count 200-1-2- 2,5  
Careline Fionia Nordic Stand up165-1-2-3  
Careline Lalandia Nordic 165-3-4

The hoists comply with essential requirements in annex 1.

The hoists also comply with the requirements described in:

**Requirements for mobile hoists edition no. 89.04**

(12 27 03 mobile hoists with slings and liftingstraps.)

Described by:

Rådet for tekniske tiltak for funksjonshemmede, Norge  
Handicappinstituttet, Sverige  
Dansk Hjælpemiddel Institut, Danmark

**The hoists comply with the requirements in DS/EN ISO 10535:1998 or DS/ EN ISO 10535:2007-02-07 2. udgave**

**Hoists for the transfer of disabled persons- Requirements and test methods.**

The hoists are CE-marked that comply after Declaration of CE conformity 02-01-2001. Signed Torben K. Petersen.

**Comply to Class I**

**Faaborg Rehab Technic ApS.**

Smedemestervej 9

5600 Faaborg

---

Niels Erik Johansen

*FAABORG REHAB TECHNIC ApS.DK-5600 Faaborg*

*TLF. (+ 45) 70205553*

*WWW.faaborg-rehab.dk E-mail [Office@faaborg-rehab.dk](mailto:Office@faaborg-rehab.dk)*

*CVR nr. 12547285*