

Applicant J.K. Medico ApS
Dam Enge 30
DK-3660 Stenløse
DENMARK

Manufacturer J.K. Medico ApS
Dam Enge 30
DK-3660 Stenløse
DENMARK

Product Battery charger

Model CCC410S
CCC408S
CCC406S

Brand Name /
Trade mark



Additional information Battery charger, IP54. Primary supply cable H05VV-F 2x0,75 mm² with non-rewirable two-pole plug Class II (Standard Sheet XVII), IP44. Secondary output cord with plug, IPXX.

IP classification IP54


The above product is tested according to the following standards EN 60529 (1991) + A1 (2000)
Degree of protection provided by enclosures (IP Code)

IP54

Nemko confirms that one sample of the mentioned product has been tested and found to be in conformity with the requirement for the degree of protection as specified above.

This statement gives no right to use of Nemko's name or certification mark on the products.

Dato 2014-09-30



Tested by: Finn-Tore Jørgensen



Approved by: Espen Eriksen