



DECLARATION OF CONFORMITY

WinnCare Nordic ApS. Taarnborgvej 12C, 4220 Korsoer, Denmark. SRN: DK-MF-000002147., as the manufacturer of following product/equipment declare compliance with the applicable provisions of Regulation (EU) 2017/745 of 05 April 2017 and the following standards: EN ISO 10535:2007 (EN 60601-1:2006, EN 60601-1-2:2015, EN ISO 14971:2019, EN ISO 13485:2016):

Brand:	WINCARE
Tradename:	Ergosling
Model/Type:	- Universal Basic - Universal Basic with head support - Universal Basic, mesh - Universal Basic with head support, mesh
From serial number:	100001 - 100002
Product classification:	ISO 9999:2016
- Product group:	12 36 21 Body-support units for hoists
- Usage:	Reusable medical device
- Sterility:	Nonsterile
Classification:	MDR Class 1 in accordance with Rule 1 / 13 of Annex VIII of Regulation (EU) 2017/745
Basic UDI-DI:	5740008105011TA
UDI number:	05740008100719, 05740008100726, 05740008100733 05740008100740, 05740008100757, 05740008100764 05740008100771, 05740008100788, 05740008100795 05740008100801, 05740008100818, 05740008100825 05740008100832, 05740008100849, 05740008100856 05740008100863, 05740008100870, 05740008100887 05740008100894, 05740008100900, 05740008100917 05740008100924, 05740008100931, 05740008100948

Korsoer, 05 February 2024

Place and date

Richard Christiansen, Technical director,
WinnCare Nordic ApS.

WinnCare Nordic ApS, Taarnborgvej 12C, 4220 Korsoer, Denmark

