



# CENTRE FOR TESTING AND CERTIFICATION - MECH-TEST

## Mechanical Laboratory

05-077 Warszawa-Wesoła, ul. Klonowa 22

tel.: +48 603 23-26-45, e-mail: cbc.mech.test@gmail.com, www.cbc.org.pl

Date 14.07.2021

### TEST REPORT NO. **CBC-163/2021**

Page 1 of 4

**Subject of testing:** *Walking aids with built-in handgrips and three or more legs of which two or more are having wheels, which provide support whilst walking* **Classification according to** PN-EN ISO 9999:2017-02 : 12 06 06

**Type / Model:** *JAGUAR SUPREME indoor rollator*

**Nr kat.:** --  
**REF:** 312415  
**SN:** 0001

**Manufacturer:** *MOBILEX A/S  
Grønlandsvej 5  
DK – 8660 Skanderborg*

**Number of specimens:** 1

**Applicant:** *A-Net s.c.  
93-469 Łódź,  
ul. Łaskowice174*


**Kind of testing** *Mechanical testing for conformity with PN-EN ISO 11199-2 : 2005*

**Test started:** 7.07.2021

**Test finished:** 14.07.2021

Approved by:

DYREKTOR

  
mgr inż. Andrzej Tkaczyk

**Special comments / enclosures:**

Copyright © 2012 by Centre for Testing and Certification (applicable to report form)

Test results refer only to tested units.

Test results reported here are not applicable to the further modifications of the product affecting its structure, material or technology.

This test report shall be neither copied differently as in the whole nor be published without written consent of the Laboratory.



CHARACTERISTIC OF PRODUCT

<b>Name :</b> JAGUAR SUPREME	<b>Dimension of rollator:</b> --
<b>SN:</b> 0001	<b>REF:</b> 312415
<b>Maximum permissible user mass:</b> 100 kg	<b>Mass of rollator:</b> 6,07 kg

Description		Comments	
Elements/parameters/materials/dimensions			
Dimensions of walking rollator (fig. 2 PN-EN ISO 11199-2)	Distance between handgrips (dimension 2)	230 mm	
	Angle between of handgrip axis and direction of movement ( $\alpha$ )	80°	
	Height of rollator (dimension 6)	834 mm	min.
		948 mm	max.
	Width of rollator (dimension 5)	569 mm	
	Turning width (dimension 1)	806 mm	
Length of rollator (dimension 4)	666 mm		
Dimensions of folded rollator (mm)		930 x 569 x 285	
Fig. 3	Handgrip - diameter	29 mm	
	Handgrip - length	142 mm	
Wheels of rollator	Front wheels - quantity	2	castor wheels
	Front wheels - diameter	177 mm	
	Front wheels - width	35/32 mm	
	Front wheels - brake	none	
	Rear wheels - quantity	2	
	Rear wheels - diameter	177 mm	
	Rear wheels - width	35/32 mm	
	Rear wheels - brake	Included	
Tip	Diameter		
	Material	Not any	
	Colour		
Material of rollator (fig. 1)	Front legs	Aluminum,	
	Bracing member (no. 8)	Steel,	
	Rear legs	Hard plastic,	
	Height adjusting device (no. 4)	Bolts, nuts	
	Handgrip (no 5),	Hard plastic	
	Brake elements		

PHOTO OF PRODUCT



**Mobilex** Jaguar SUPREME indoor rollator  
a better life in motion

MOBILEX A/S  
Grønlandsvej 5  
DK - 8660 Skanderborg  
Tel: +45 87 93 22 20  
www.mobilex.dk

REF 312415

≤ 100 kg

54.5 cm


≤ 5 kg

2020.12.07

0001

SN

5 740001 437034



## RESULT OF TESTS ACCORDING TO PN-EN ISO 11199-2:2005

Requirements according to clause	Test method according to clause	Checked characteristics/assemblies/parameters		Real value	Test result	Comments	
4.1	Measur.	Manoeuvrability		$\phi$ 177 mm width 32mm Conf.	Pos.	$\phi$ front wheels $\geq 75$ mm outdoor intended rollator: $\phi$ front wheels $\geq 180$ mm width of wheels $\geq 28$ mm	
4.2	5.3	Forward-direction stability		21,4 <sup>0</sup> Conf.	Pos.	Stability required $\geq 10^\circ$	
4.2	5.4	Backward-direction stability		12,4 <sup>0</sup> Conf.	Pos.	Stability required $\geq 7^\circ$	
4.2	5.5	Sideway-direction stability		6,8 <sup>0</sup> Conf.	Pos.	Stability required $\geq 3,5^\circ$	
4.2	5.6	Stability – with loaded basket, bag (5kg), drip, oxygen cylinder		forwards	20,8 <sup>0</sup> Conf.	Pos.	Stability required $\geq 10^\circ$
				backwards	14,5 <sup>0</sup> Conf.	Pos.	Stability required $\geq 7^\circ$
				side	8,2 <sup>0</sup> Conf.	Pos.	Stability required $\geq 3,5^\circ$
4.3	V/I	Brakes	Servicing facility during rollator motion with more than 2 wheels	Conf.	Pos.		
	V/I		Parking brakes in rollator with more than 2 wheels and resting seat or intended for outdoor use	Conf.	Pos.		
	5.7.1.1		Brake grip distance (fig. 4, dimension 1)	72 mm Conf.	Pos.	$\leq 75$ mm	
	5.7.1		Running brake effectiveness	Conf.	Pos.	Movement of rollator $\leq 10$ mm in 1 minute	
	Measur.		Force to set parking brake	60 N Conf.	Pos.	$\leq 60$ N	
	Measur.		Force to release parking brake	--	N/A	$\leq 40$ N	
	5.7.2		Parking brake effectiveness	Conf.	Pos.	Movement of rollator $\leq 10$ mm in 1 minute	
	V/I		Possibility to compensate brake wear	Conf.	Pos.		
V/I	Brake not adversely affected by folding, unfolding or adjusting actions of rollator	Conf.	Pos.				
4.4	Measur. V/I	Handgrip		29 mm Conf.	Pos.	Width of handgrip $\geq 20$ mm and $\leq 50$ mm	
4.5	Measur. V/I	Leg section and tip		--	N/A	$\phi$ tip $\geq 35$ mm (tested rollator is equipped in four wheels)	
4.6	5.10	Resting seat – static loading durability		--	N/A		
4.7	5.12	Mechanical durability	Fatigue test	Conf.	Pos.	200 000 cycles with load. 80 kg $\pm$ 2%, f=1Hz	
			Static loading test	Conf.	Pos.	loading 120kg $\pm$ 2%, 5sek. NOTE 1	
4.8	V/I	Adjusting devices		Conf.	Pos.		
4.9	5.14	Folding mechanism		Conf.	Pos.		
4.11	ISO 10993-1	Materials and finish	Biocompatibility of material with human body	--	N/T		
	V/I		Free of discolouring of skin or clothing in contact with rollator materials	Conf.	Pos.		
	V/I		Burrs, shar edges, projections	Conf.	Pos.		
Marking and labelling of product							
6.2	V/I	a) Maximum user mass		Included	Pos.		
		b) Maximum safe working load (SWL) to be marked on accessories		Included	Pos.		
		c) Maximum allowed angle between the longitudinal centreline of the handle and the direction of motion, if the handles are sideways adjustable		--	N/A		
		d) Manufacturer's name or trade name and address		Included	Pos.		
		e) Manufacturer's model identification name and/or number		Included	Pos.		
		f) Month and year of manufacture		Included	Pos.		
		g) Maximum extension of the height adjustment, marked on the adjusting members		Included	Pos.		
		h) Maximum width of the rollator		Included	Pos.		
		i) Rollator intended for outdoor/indoor use		Included	Pos.		
4.10	V/I	Warning showing allowed angle between handle axis and direction of movement or physical stop of angle adjusting		--	N/A		

## Contents of user manual and/or assembly manual or clear and indelible marking of product

6.3	V/I	a) Maximum rollator height	Included	Pos.	
		b) Minimum rollator height	Included	Pos.	
		c) maintenance and cleaning instructions, including a description of the method and suitable cleaning agents and any precautions needed to avoid corrosion and/or ageing of the materials used in construction of the rollator	Included	Pos.	
		d) Instructions for assembly, adjustment of all kinds, folding and unfolding	Included	Pos.	
		e) Warnings and advice about precautions relating to safe distances between moving and stationary parts (see EN 12182, Clauses 12 and 13, for guidance)	Included	Pos.	
		f) Maximum safe working load (SWL) for load carrying accessories such as basket, tray, shopping bag, etc.	Included	Pos.	
4.10	V/I	Warning in user manual on consequences of such an adjustment of angle between handle longitudinal axis and direction of movement outside allowed value (when handles are adjustable aside).	--	N/A	

## TEST CONDITIONS

Ambient temperature	19°C	Required temperature 21°C ±5°C
Relative humidity of air:	55 %	Not required
<i>Comments:</i>		
All tests performed with maximum height adjustment of rollator.		
All tests performed in the least stable position of self-adjusting wheels.		
Tests performed with handles positioned at their maximum (allowed) angle to the direction of motion (when adjustment is possible).		
Sequence of tests: stability test, static loading test, fatigue test.		
One rollator was tested.		
During visual inspection before testing any visible defects that could have influence on test results were not stated.		

Pos. – positive; Neg – negative; N/T – not tested; N/A – not applicable; N/R – not required, N/O – not occurred, V/I.- visual inspection, Conf.- conformed.

NOTE 1: Deformation – 35 mm, elastic deformation - 34 mm, permanent deformation – 1 mm (0,1%).

## CONCLUSIONS:

Testing object **conforms** with requirements of PN-EN ISO 11199-2 : 2005 in scope of mechanical testing ordered by client, excluding biocompatibility tests of material with human body according to PN-EN ISO 10993-1:2010 .

---- END ----