



# CENTRE FOR TESTING AND CERTIFICATION - MECH-TEST

## Mechanical Laboratory

05-077 Warszawa-Wesoła, ul. Klonowa 22  
tel.: +48 603 23-26-45, e-mail: cbc.mech.test@gmail.com, www.cbc.org.pl

Date 20.05.2013


### TEST REPORT NO. **CBC-097/2013**

Page 1 of 6

<b>Subject of testing:</b>	<i>Rubber Ramp</i>	<b>Classification according to</b>	<b>PN-EN ISO 9999:2011 : 18 30 15</b>
<b>Type / Model:</b>	<i>48x250mm</i>	<b>Factory ref. no.:</b>	<i>306482</i>
<b>Manufacturer:</b>	<i>MOBILEX A/S Nørskovvej 1 DK – 8660 Skanderborg</i>	<b>Number of specimens:</b>	<i>1</i>
<b>Applicant:</b>	<i>A-Net s.c. 93-469 Łódź, ul. Łaskowice174</i>		
<b>Kind of testing</b>	<i>Testing scope according to application of Client Mechanical testing according to PN-EN 12182:2012;</i>		
<b>Test started:</b>	<i>10.05.2013</i>		
<b>Test finished:</b>	<i>20.05.2013</i>		

Approved by:

DYREKTOR

  
mgr inż. Andrzej Tkaczyk

Special comments / enclosures:

Copyright © 2012 by Centre for Testing and Certification (applicable to report form)

Test results refer only to tested units.

Test results reported here are not applicable to the further modifications of the product affecting its structure, material or technology.



This test report shall be neither copied differently as in the whole nor be published without written consent of the Laboratory.




**TESTING****NORMATIVE REFERENCES**

PN-EN 12182:2012

Technical aids for disabled persons – General requirements and test methods

*PHOTO OF PRODUCT*

one wheel load - 225kg

**CHARAKTERISTICS OF PRODUCT**

<b>Name of product:</b> <i>Rubber Ramp</i>	<i>Max. loading:</i> <b>300 kg</b>	<i>Mass of product:</i> <b>6,03 kg</b>
<i>Dimensions:</i>		
<i>Height (max.) :</i>	<b>48 mm</b>	
<i>Width (max.):</i>	<b>260 mm</b>	
<i>Length max. :</i>	<b>1000 mm</b>	
<i>of the ramp to transport:</i>	--	
<i>Material:</i>		
<b>rubber</b>		