



CENTRE FOR TESTING AND CERTIFICATION - MECH-TEST

Mechanical Laboratory

05-077 Warszawa-Wesoła, ul. Klonowa 22
tel.: +48 603 23-26-45, e-mail: cbc.mech.test@gmail.com, www.cbc.org.pl

Date 30.11.2015

TEST REPORT NO. *CBC-100/2015*

Page 1 of 10

Subject of testing:	<i>Walking aids with built-in handgrips and three or more legs of which two or more are having wheels, which provide support whilst walking</i>	Classification according to PN-EN ISO 9999:2011 : 12 06 06
Type / Model:	<i>Impala rollator For outdoor and indoor use</i>	Art. nr.: 312060 312061 312062
Manufacturer:	<i>MOBILEX A/S Nørskovvej 1 DK-8660 Skanderborg</i>	Number of specimens: 3
Applicant:	<i>A-Net s.c. 93-469 Łódź, ul. Łaskowice 174</i>	
Kind of testing	<i>Mechanical testing for conformity with PN-EN ISO 11199-2 : 2005</i>	
Test started:	<i>29.09.2015</i>	
Test finished:	<i>30.11.2015</i>	

Approved by:

DYREKTOR

mgr inż. Andrzej Tkaczyk

Special comments / enclosures:

Copyright © 2012 by Centre for Testing and Certification (applicable to report form)

Test results refer only to tested units.

Test results reported here are not applicable to the further modifications of the product affecting its structure, material or technology.


This test report shall be neither copied differently as in the whole nor be published without written consent of the Laboratory.



CHARACTERISTIC OF PRODUCT

Name : IMPALA rollator	Dimension of rollator: --
SN: --	Product code: 312060
Maximum permissible user mass: 150 kg	Mass of rollator: 7,50 kg

Description			PHOTO OF PRODUCT	
Elements/parameters/materials/dimensions	Comments			
Dimensions of walking rollator (fig. 2 PN-EN ISO 11199-2)	Distance between handgrips (dimension 2)	477mm		
	Angle between of handgrip axis and direction of movement (α)	0°		
	Height of rollator (dimension 6)	785 mm		min.
		925 mm		max.
	Width of rollator (dimension 5)	606 mm		
	Turning width (dimension 1)	872 mm		
Length of rollator (dimension 4)	697 mm			
Dimensions of folded rollator (mm)	945 x 606 x 330			
Fig. 3	Handgrip - diameter	31 mm		Anatomical handgrip
	Handgrip - length	96 mm		
Wheels of rollator	Front wheels- quantity	2		castor wheels
	Front wheels - diameter	199 mm		
	Front wheels - width	34 mm		
	Front wheels - brake	none		
	Rear wheels - quantity	2		
	Rear wheels - diameter	199mm		
	Rear wheels - width	34 mm		
Rear wheels - brake	Included			
Tip	Diameter			
	Material	Not any		
	Colour			
Material of rollator (fig. 1)	Front legs	Aluminum,		
	Bracing member (no. 8)	Steel,		
	Rear legs	Hard plastic,		
	Height adjusting device (no. 4)	Bolts, nuts		
	Handgrip (no 5), Brake elements	Hard plastic		



RESULT OF TESTS ACCORDING TO PN-EN ISO 11199-2:2005

Requirements according to clause	Test method according to clause	Checked characteristics/assemblies/parameters		Real value	Test result	Comments
4.1	Measur.	Manoeuvrability		\varnothing 199 mm width 30mm Conf.	Pos.	\varnothing front wheels ≥ 75 mm outdoor intended rollator: \varnothing front wheels ≥ 180 mm width of wheels ≥ 28 mm
4.2	5.3	Forward-direction stability		25,0° Conf.	Pos.	Stability required $\geq 10^\circ$
4.2	5.4	Backward-direction stability		11,5° Conf.	Pos.	Stability required $\geq 7^\circ$
4.2	5.5	Sideway-direction stability		4,2° Conf.	Pos.	Stability required $\geq 3,5^\circ$
4.2	5.6	Stability – with loaded basket, bag, drip, oxygen cylinder (max. load bags – 5 kg)	forwards	24,0° Conf.	Pos.	Stability required $\geq 10^\circ$
			backwards	14,5° Conf.	Pos.	Stability required $\geq 7^\circ$
			side	6,0° Conf.	Pos.	Stability required $\geq 3,5^\circ$
4.3	V/I	Brakes	Servicing facility during rollator motion with more than 2 wheels	Conf.	Pos.	
	V/I		Parking brakes in rollator with more than 2 wheels and resting seat or intended for outdoor use	Conf.	Pos.	
	5.7.1.1		Brake grip distance (fig. 4, dimension 1)	75 mm Conf.	Pos.	≤ 75 mm
	5.7.1		Running brake effectiveness	Conf.	Pos.	Movement of rollator ≤ 10 mm in 1 minute
	Measur.		Force to set parking brake	30N Conf.	Pos.	≤ 60 N
	Measur.		Force to release parking brake	10N Conf.	Pos.	≤ 40 N
	5.7.2		Parking brake effectiveness	Conf.	Pos.	Movement of rollator ≤ 10 mm in 1 minute
	V/I		Possibility to compensate brake wear	Conf.	Pos.	
V/I	Brake not adversely affected by folding, unfolding or adjusting actions of rollator	Conf.	Pos.			
4.4	Measur. V/I	Handgrip		31 mm Conf.	Pos.	Width of handgrip ≥ 20 mm and ≤ 50 mm
4.5	Measur. V/I	Leg section and tip		--	N/A	\varnothing tip ≥ 35 mm (tested rollator is equipped in four wheels)
4.6	5.10	Resting seat – static loading durability		Conf.	Pos.	1 minute under load 1,2 x user's weight $\pm 2\%$ (180 kg)
4.7	5.12	Mechanical durability	Fatigue test	Conf.	Pos.	200 000 cycles with load. 120 kg $\pm 2\%$, f=1Hz
4.7	5.11		Static loading test	Conf.	Pos.	loading 180 kg $\pm 2\%$, 5sek.
4.8	V/I	Adjusting devices		Conf.	Pos.	
4.9	5.14	Folding mechanism		Conf.	Pos.	
4.11	ISO 10993-1	Material s and finish	Biocompatibility of material with human body	--	N/T	
	V/I		Free of discolouring of skin or clothing in contact with rollator materials	Conf.	Pos.	
	V/I		Burrs, shar edges, projections	Conf.	Pos.	
Marking and labelling of product						
6.2	V/I	a) Maximum user mass		Included	Pos.	
		b) Maximum safe working load (SWL) to be marked on accessories		Included	Pos.	
		c) Maximum allowed angle between the longitudinal centreline of the handle and the direction of motion, if the handles are sideways adjustable		--	N/A	angle between direction of motion and longitudinal axis of handgrip not adjustable
		d) Manufacturer's name or trade name and address		Included	Pos.	
		e) Manufacturer's model identification name and/or number		Included	Pos.	
		f) Month and year of manufacture		Included	Pos.	
		g) Maximum extension of the height adjustment, marked on the adjusting members		Included	Pos.	
		h) Maximum width of the rollator		Included	Pos.	
		i) Rollator intended for outdoor/indoor use		Included	Pos.	
4.10	V/I	Warning showing allowed angle between handle axis and direction of movement or physical stop of angle adjusting		--	N/A	angle between direction of motion and longitudinal axis of handgrip not adjustable

Contents of user manual and/or assembly manual or clear and indelible marking of product					
6.3	V/I	a) Maximum rollator height	--	N/T	
		b) Minimum rollator height	--	N/T	
		c) maintenance and cleaning instructions, including a description of the method and suitable cleaning agents and any precautions needed to avoid corrosion and/or ageing of the materials used in construction of the rollator	--	N/T	
		d) Instructions for assembly, adjustment of all kinds, folding and unfolding	--	N/T	
		e) Warnings and advice about precautions relating to safe distances between moving and stationary parts (see EN 12182, Clauses 12 and 13, for guidance)	--	N/T	
		f) Maximum safe working load (SWL) for load carrying accessories such as basket, tray, shopping bag, etc.	--	N/T	
4.10	V/I	Warning in user manual on consequences of such an adjustment of angle between handle longitudinal axis and direction of movement outside allowed value (when handles are adjustable aside).	--	N/A	angle between direction of motion and longitudinal axis of handgrip not adjustable
TEST CONDITIONS					
Ambient temperature			21°C	Required temperature 21°C ±5°C	
Relative humidity of air:			60 %	Not required	
Comments:					
All tests performed with maximum height adjustment of rollator.					
All tests performed in the least stable position of self-adjusting wheels.					
Tests performed with handles positioned at their maximum (allowed) angle to the direction of motion (when adjustment is possible).					
Sequence of tests: stability test, static loading test, fatigue test.					
One rollator was tested.					
During visual inspection before testing any visible defects that could have influence on test results were not stated.					

Pos. – positive; Neg – negative; N/T – not tested; N/A – not applicable; N/R – not required, N/O – not occurred, V/I.- visual inspection, Conf.- conformed.

NOTE 1: An additional fatigue tests on the two-drum stand with obstacles of a height of 12mm, according to ISO 7176-8: 2002, p. 10.4.2., was performed.

- ▶ The handles were loaded with a load of 120 kg. After 15 000 cycles, the rollator was not damaged. The test result was positive.
- ▶ The seat was loaded with a load of 100 kg. After a further 35 000 cycles, the rollator was not damaged. The test result positive

NOTE 2: Additional Drop Pouch tests were performed. The seat was loaded with a mass of 100 kg.

Rear of the rollator was dropped (on wheels) 5 times from a height of 100mm. Rollator was not damaged. The test result positive.

CONCLUSIONS:

Testing object **conforms** with requirements of PN-EN ISO 11199-2 : 2005, in scope of mechanical testing ordered by client, excluding biocompatibility tests of material with human body according to PN-EN ISO 10993-1:2010

CHARACTERISTIC OF PRODUCT

Name : IMPALA rollator	Dimension of rollator: --
SN: --	Product code: 312061
Maximum permissible user mass: 150 kg	Mass of rollator: 8,21kg

Description		PHOTO OF PRODUCT	
Elements/parameters/materials/dimensions		Comments	
Dimensions of walking rollator (fig. 2 PN-EN ISO 11199-2)	Distance between handgrips (dimension 2)	490 mm	
	Angle between of handgrip axis and direction of movement (α)	0°	
	Height of rollator (dimension 6)	900 mm	min.
		1045 mm	max.
	Width of rollator (dimension 5)	635 mm	
	Turning width (dimension 1)	910 mm	
	Length of rollator (dimension 4)	745mm	
Dimensions of folded rollator (mm)		1045 x 635 x 325	
Fig. 3	Handgrip - diameter	31 mm	Anatomical handgrip
	Handgrip - length	96 mm	
Wheels of rollator	Front wheels - quantity	2	castor wheels
	Front wheels - diameter	190 mm	
	Front wheels - width	35 mm	
	Front wheels - brake	none	
	Rear wheels - quantity	2	
	Rear wheels - diameter	190mm	
	Rear wheels - width	35 mm	
Rear wheels - brake	Included		
Tip	Diameter		
	Material	Not any	
	Colour		
Material of rollator (fig. 1)	Front legs	Aluminum,	
	Bracing member (no. 8)	Steel,	
	Rear legs	Hard plastic,	
	Height adjusting device (no. 4)	Bolts, nuts	
	Handgrip (no 5), Brake elements	Hard plastic	



RESULT OF TESTS ACCORDING TO PN-EN ISO 11199-2:2005

Requirements according to clause	Test method according to clause	Checked characteristics/assemblies/parameters		Real value	Test result	Comments
4.1	Measur.	Manoeuvrability		\varnothing 190 mm width 35mm Conf.	Pos.	\varnothing front wheels ≥ 75 mm outdoor intended rollator: \varnothing front wheels ≥ 180 mm width of wheels ≥ 28 mm
4.2	5.3	Forward-direction stability		24,0 ⁰ Conf.	Pos.	Stability required $\geq 10^\circ$
4.2	5.4	Backward-direction stability		12,5 ⁰ Conf.	Pos.	Stability required $\geq 7^\circ$
4.2	5.5	Sideway-direction stability		5,0 ⁰ Conf.	Pos.	Stability required $\geq 3,5^\circ$
4.2	5.6	Stability – with loaded basket, bag, drip, oxygen cylinder (max. load bags – 5 kg)	forwards	24,7 ⁰ Conf.	Pos.	Stability required $\geq 10^\circ$
			backwards	14,5 ⁰ Conf.	Pos.	Stability required $\geq 7^\circ$
			side	7,0 ⁰ Conf.	Pos.	Stability required $\geq 3,5^\circ$
4.3	V/I	Brakes	Servicing facility during rollator motion with more than 2 wheels	Conf.	Pos.	
	V/I		Parking brakes in rollator with more than 2 wheels and resting seat or intended for outdoor use	Conf.	Pos.	
	5.7.1.1		Brake grip distance (fig. 4, dimension 1)	75 mm Conf.	Pos.	≤ 75 mm
	5.7.1		Running brake effectiveness	Conf.	Pos.	Movement of rollator ≤ 10 mm in 1 minute
	Measur.		Force to set parking brake	30N Conf.	Pos.	≤ 60 N
	Measur.		Force to release parking brake	10N Conf.	Pos.	≤ 40 N
	5.7.2		Parking brake effectiveness	Conf.	Pos.	Movement of rollator ≤ 10 mm in 1 minute
	V/I		Possibility to compensate brake wear	Conf.	Pos.	
V/I	Brake not adversely affected by folding, unfolding or adjusting actions of rollator	Conf.	Pos.			
4.4	Measur. V/I	Handgrip		31 mm Conf.	Pos.	Width of handgrip ≥ 20 mm and ≤ 50 mm
4.5	Measur. V/I	Leg section and tip		--	N/A	\varnothing tip ≥ 35 mm (tested rollator is equipped in four wheels)
4.6	5.10	Resting seat – static loading durability		Conf.	Pos.	1 minute under load 1,2 x user's weight $\pm 2\%$ (180 kg)
4.7	5.12	Mechanical durability	Fatigue test	Conf.	Pos.	200 000 cycles with load. 120 kg $\pm 2\%$, f=1Hz
4.7	5.11		Static loading test	Conf.	Pos.	loading 180 kg $\pm 2\%$, 5sek.
4.8	V/I	Adjusting devices		Conf.	Pos.	
4.9	5.14	Folding mechanism		Conf.	Pos.	
4.11	ISO 10993-1	Material s and finish	Biocompatibility of material with human body	--	N/T	
	V/I		Free of discolouring of skin or clothing in contact with rollator materials	Conf.	Pos.	
	V/I		Burrs, shar edges, projections	Conf.	Pos.	
Marking and labelling of product						
6.2	V/I	a) Maximum user mass		Included	Pos.	
		b) Maximum safe working load (SWL) to be marked on accessories		Included	Pos.	
		c) Maximum allowed angle between the longitudinal centreline of the handle and the direction of motion, if the handles are sideways adjustable		--	N/A	angle between direction of motion and longitudinal axis of handgrip not adjustable
		d) Manufacturer's name or trade name and address		Included	Pos.	
		e) Manufacturer's model identification name and/or number		Included	Pos.	
		f) Month and year of manufacture		Included	Pos.	
		g) Maximum extension of the height adjustment, marked on the adjusting members		Included	Pos.	
		h) Maximum width of the rollator		Included	Pos.	
		i) Rollator intended for outdoor/indoor use		Included	Pos.	
4.10	V/I	Warning showing allowed angle between handle axis and direction of movement or physical stop of angle adjusting		--	N/A	angle between direction of motion and longitudinal axis of handgrip not adjustable

Contents of user manual and/or assembly manual or clear and indelible marking of product					
6.3	V/I	a) Maximum rollator height	--	N/T	
		b) Minimum rollator height	--	N/T	
		c) maintenance and cleaning instructions, including a description of the method and suitable cleaning agents and any precautions needed to avoid corrosion and/or ageing of the materials used in construction of the rollator	--	N/T	
		d) Instructions for assembly, adjustment of all kinds, folding and unfolding	--	N/T	
		e) Warnings and advice about precautions relating to safe distances between moving and stationary parts (see EN 12182, Clauses 12 and 13, for guidance)	--	N/T	
		f) Maximum safe working load (SWL) for load carrying accessories such as basket, tray, shopping bag, etc.	--	N/T	
4.10	V/I	Warning in user manual on consequences of such an adjustment of angle between handle longitudinal axis and direction of movement outside allowed value (when handles are adjustable aside).	--	N/A	angle between direction of motion and longitudinal axis of handgrip not adjustable
TEST CONDITIONS					
Ambient temperature			21°C	Required temperature 21°C ±5°C	
Relative humidity of air:			60 %	Not required	
Comments:					
All tests performed with maximum height adjustment of rollator.					
All tests performed in the least stable position of self-adjusting wheels.					
Tests performed with handles positioned at their maximum (allowed) angle to the direction of motion (when adjustment is possible).					
Sequence of tests: stability test, static loading test, fatigue test.					
One rollator was tested.					
During visual inspection before testing any visible defects that could have influence on test results were not stated.					

Pos. – positive; Neg – negative; N/T – not tested; N/A – not applicable; N/R – not required, N/O – not occurred, V/I.- visual inspection, Conf.- conformed.

NOTE 1: An additional fatigue tests on the two-drum stand with obstacles of a height of 12mm, according to ISO 7176-8: 2002, p. 10.4.2., was performed.

- ▶ The handles were loaded with a load of 120 kg. After 15 000 cycles, the rollator was not damaged. The test result was positive.
- ▶ The seat was loaded with a load of 100 kg. After a further 35 000 cycles, the rollator was not damaged. The test result positive.

NOTE 2: Additional Drop Pouch tests were performed. The seat was loaded with a mass of 100 kg. Rear of the rollator was dropped (on wheels) 5 times from a height of 100mm. Rollator was not damaged. The test result positive.

CONCLUSIONS:

Testing object **conforms** with requirements of PN-EN ISO 11199-2 : 2005, in scope of mechanical testing ordered by client, excluding biocompatibility tests of material with human body according to PN-EN ISO 10993-1:2010

CHARACTERISTIC OF PRODUCT

Name : IMPALA rollator	Dimension of rollator: --
SN: --	Product code: 312062
Maximum permissible user mass: 150 kg	Mass of rollator: 7,22 kg

Description			Comments	PHOTO OF PRODUCT	
Elements/parameters/materials/dimensions					
Dimensions of walking rollator (fig. 2 PN-EN ISO 11199-2)	Distance between handgrips (dimension 2)	530 mm			
	Angle between of handgrip axis and direction of movement (α)	0°			
	Height of rollator (dimension 6)	785 mm	min.		
		925 mm	max.		
	Width of rollator (dimension 5)	666 mm			
	Turning width (dimension 1)	902mm			
	Length of rollator (dimension 4)	697 mm			
Dimensions of folded rollator (mm)		845 x 666 x 330			
Fig. 3	Handgrip - diameter	31 mm	Anatomical handgrip		
	Handgrip - length	96 mm			
Wheels of rollator	Front wheels - quantity	2	castor wheels		
	Front wheels - diameter	200 mm			
	Front wheels - width	29 mm			
	Front wheels - brake	none			
	Rear wheels - quantity	2			
	Rear wheels - diameter	200 mm			
	Rear wheels - width	29 mm			
Rear wheels - brake	Included				
Tip	Diameter				
	Material	Not any			
	Colour				
Material of rollator (fig. 1)	Front legs	Aluminum,			
	Bracing member (no. 8)	Steel,			
	Rear legs	Hard plastic,			
	Height adjusting device (no. 4)	Bolts, nuts			
	Handgrip (no 5), Brake elements	Hard plastic			



RESULT OF TESTS ACCORDING TO PN-EN ISO 11199-2:2005

Requirements according to clause	Test method according to clause	Checked characteristics/assemblies/parameters		Real value	Test result	Comments
4.1	Measur.	Manoeuvrability		\varnothing 200 mm width 28 mm Conf.	Pos.	\varnothing front wheels ≥ 75 mm outdoor intended rollator: \varnothing front wheels ≥ 180 mm width of wheels ≥ 28 mm
4.2	5.3	Forward-direction stability		26,0 ⁰ Conf.	Pos.	Stability required $\geq 10^\circ$
4.2	5.4	Backward-direction stability		10,9 ⁰ Conf.	Pos.	Stability required $\geq 7^\circ$
4.2	5.5	Sideway-direction stability		4,7 ⁰ Conf.	Pos.	Stability required $\geq 3,5^\circ$
4.2	5.6	Stability – with loaded basket, bag, drip, oxygen cylinder (max. load bags – 5 kg)	forwards	25,0 ⁰ Conf.	Pos.	Stability required $\geq 10^\circ$
			backwards	14,3 ⁰ Conf.	Pos.	Stability required $\geq 7^\circ$
			side	6,7 ⁰ Conf.	Pos.	Stability required $\geq 3,5^\circ$
4.3	V/I	Brakes	Servicing facility during rollator motion with more than 2 wheels	Conf.	Pos.	
	V/I		Parking brakes in rollator with more than 2 wheels and resting seat or intended for outdoor use	Conf.	Pos.	
	5.7.1.1		Brake grip distance (fig. 4, dimension 1)	75 mm Conf.	Pos.	≤ 75 mm
	5.7.1		Running brake effectiveness	Conf.	Pos.	Movement of rollator ≤ 10 mm in 1 minute
	Measur.		Force to set parking brake	30N Conf.	Pos.	≤ 60 N
	Measur.		Force to release parking brake	10N Conf.	Pos.	≤ 40 N
	5.7.2		Parking brake effectiveness	Conf.	Pos.	Movement of rollator ≤ 10 mm in 1 minute
	V/I		Possibility to compensate brake wear	Conf.	Pos.	
V/I	Brake not adversely affected by folding, unfolding or adjusting actions of rollator	Conf.	Pos.			
4.4	Measur. V/I	Handgrip		31 mm Conf.	Pos.	Width of handgrip ≥ 20 mm and ≤ 50 mm
4.5	Measur. V/I	Leg section and tip		--	N/A	\varnothing tip ≥ 35 mm (tested rollator is equipped in four wheels)
4.6	5.10	Resting seat – static loading durability		Conf.	Pos.	1 minute under load 1,2 x user's weight $\pm 2\%$ (180 kg)
4.7	5.12	Mechanical durability	Fatigue test	Conf.	Pos.	200 000 cycles with load. 120 kg $\pm 2\%$, f=1Hz
4.7	5.11		Static loading test	Conf.	Pos.	loading 180 kg $\pm 2\%$, 5sek.
4.8	V/I	Adjusting devices		Conf.	Pos.	
4.9	5.14	Folding mechanism		Conf.	Pos.	
4.11	ISO 10993-1	Material s and finish	Biocompatibility of material with human body	--	N/T	
	V/I		Free of discolouring of skin or clothing in contact with rollator materials	Conf.	Pos.	
	V/I		Burrs, shar edges, projections	Conf.	Pos.	
Marking and labelling of product						
6.2	V/I	a) Maximum user mass		Included	Pos.	
		b) Maximum safe working load (SWL) to be marked on accessories		Included	Pos.	
		c) Maximum allowed angle between the longitudinal centreline of the handle and the direction of motion, if the handles are sideways adjustable		--	N/A	angle between direction of motion and longitudinal axis of handgrip not adjustable
		d) Manufacturer's name or trade name and address		Included	Pos.	
		e) Manufacturer's model identification name and/or number		Included	Pos.	
		f) Month and year of manufacture		Included	Pos.	
		g) Maximum extension of the height adjustment, marked on the adjusting members		Included	Pos.	
		h) Maximum width of the rollator		Included	Pos.	
		i) Rollator intended for outdoor/indoor use		Included	Pos.	
4.10	V/I	Warning showing allowed angle between handle axis and direction of movement or physical stop of angle adjusting		--	N/A	angle between direction of motion and longitudinal axis of handgrip not adjustable

Contents of user manual and/or assembly manual or clear and indelible marking of product

6.3	V/I	a) Maximum rollator height	--	N/T	
		b) Minimum rollator height	--	N/T	
		c) maintenance and cleaning instructions, including a description of the method and suitable cleaning agents and any precautions needed to avoid corrosion and/or ageing of the materials used in construction of the rollator	--	N/T	
		d) Instructions for assembly; adjustment of all kinds, folding and unfolding	--	N/T	
		e) Warnings and advice about precautions relating to safe distances between moving and stationary parts (see EN 12182, Clauses 12 and 13, for guidance)	--	N/T	
		f) Maximum safe working load (SWL) for load carrying accessories such as basket, tray, shopping bag, etc.	--	N/T	
4.10	V/I	Warning in user manual on consequences of such an adjustment of angle between handle longitudinal axis and direction of movement outside allowed value (when handles are adjustable aside).	--	N/A	angle between direction of motion and longitudinal axis of handgrip not adjustable

TEST CONDITIONS

Ambient temperature	21°C	Required temperature 21°C ±5°C
Relative humidity of air:	60 %	Not required
Comments:		
All tests performed with maximum height adjustment of rollator.		
All tests performed in the least stable position of self-adjusting wheels.		
Tests performed with handles positioned at their maximum (allowed) angle to the direction of motion (when adjustment is possible).		
Sequence of tests: stability test, static loading test, fatigue test.		
One rollator was tested.		
During visual inspection before testing any visible defects that could have influence on test results were not stated.		

Pos. – positive; Neg – negative; N/T – not tested; N/A – not applicable; N/R – not required, N/O – not occurred, V/I.- visual inspection, Conf.- conformed.

NOTE 1: An additional fatigue tests on the two-drum stand with obstacles of a height of 12mm, according to ISO 7176-8: 2002, p. 10.4.2., was performed.

- ▶ The handles were loaded with a load of 120 kg. After 15 000 cycles, the rollator was not damaged. The test result was positive.
- ▶ The seat was loaded with a load of 100 kg. After a further 35 000 cycles, the rollator was not damaged. The test result positive

NOTE 2: Additional Drop Pouch tests were performed. The seat was loaded with a mass of 100 kg. Rear of the rollator was dropped (on wheels) 5 times from a height of 100mm. Rollator was not damaged. The test result positive.

CONCLUSIONS:

Testing object **conforms** with requirements of PN-EN ISO 11199-2 : 2005, in scope of mechanical testing ordered by client, excluding biocompatibility tests of material with human body according to PN-EN ISO 10993-1:2010

---- END ----