



CENTRE FOR TESTING AND CERTIFICATION - MECH-TEST

Mechanical Laboratory

05-077 Warszawa-Wesoła, ul. Klonowa 22
tel.: +48 603 23-26-45, e-mail: cbc.mech.test@gmail.com, www.cbc.org.pl

Date 5.01.2016

TEST REPORT NO. **CBC-147/2015**

Page 1 of 4

Subject of testing:	<i>Walking aids with built-in handgrips and three or more legs of which two or more are having wheels, which provide support whilst walking</i>	Classification according to PN-EN ISO 9999:2011 : 12 06 06
Type / Model:	<i>Explorer Rollator MEDIUM</i>	Nr kat.: ROL00EM SN: 0002
Manufacturer:	<i>REHASENSE Sp. z o.o. Sulejowska 45 97-300 Piotrków Trybunalski</i>	Number of specimens: 1
Applicant:	<i>A-Net s.c. 93-469 Łódź, ul. Łaskowice174</i>	
Kind of testing	<i>Mechanical testing for conformity with PN-EN ISO 11199-2 : 2005 excluding clause 4.10, 6.2, 6.3</i>	
Test started:	17.12.2015	
Test finished:	5.01.2016	

Approved by:

DYREKTOR


mgr inż. Andrzej Tkaczyk

Special comments / enclosures:

Copyright © 2012 by Centre for Testing and Certification (applicable to report form)

Test results refer only to tested units.

Test results reported here are not applicable to the further modifications of the product affecting its structure, material or technology.

This test report shall be neither copied differently as in the whole nor be published without written consent of the Laboratory.





CHARACTERISTIC OF PRODUCT

Name : <i>MEDIUM</i>	Dimension of rollator: --
Nr. kat: <i>ROL00EM</i>	Product code: --
Maximum permissible user mass: <i>200 kg</i>	Mass of rollator: <i>9,095 kg</i>

Description		Comments	
Elements/parameters/materials/dimensions			
Dimensions of walking rollator (fig. 2 PN-EN ISO 11199-2)	Distance between handgrips (dimension 2)	550 mm	
	Angle between of handgrip axis and direction of movement (α)	0°	
	Height of rollator (dimension 6)	770 mm	min.
		903 mm	max.
	Width of rollator (dimension 5)	692 mm	
	Turning width (dimension 1)	950 mm	
Length of rollator (dimension 4)	767 mm		
Dimensions of folded rollator (mm)		767 x 835 x 253	
Fig. 3	Handgrip - diameter	37 mm	Anatomical handgrip
	Handgrip - length	120 mm	
Wheels of rollator	Front wheels - quantity	2	castor wheels
	Front wheels - diameter	282 mm	
	Front wheels - width	43 mm	
	Front wheels - brake	none	
	Rear wheels - quantity	2	
	Rear wheels - diameter	232 mm	
	Rear wheels - width	44 mm	
Rear wheels - brake	Included		
Tip	Diameter		
	Material	Not any	
	Colour		
Material of rollator (fig. 1)	Front legs	Aluminum,	
	Bracing member (no. 8)	steel, plastic,	
	Rear legs	Bolts, nuts	
	Height adjusting device (no. 4)		
	Handgrip (no 5), Brake elements	plastic	

PHOTO OF PRODUCT



RESULT OF TESTS ACCORDING TO PN-EN ISO 11199-2:2005

Requirements according to clause	Test method according to clause	Checked characteristics/assemblies/parameters		Real value	Test result	Comments
4.1	Measur.	Manoeuvrability		ϕ 282 mm width 41 mm Conf.	Pos.	ϕ front wheels ≥ 75 mm <u>outdoor intended rollator:</u> ϕ front wheels ≥ 180 mm width of wheels ≥ 22 mm
4.2	5.3	Forward-direction stability		20,7 ⁰ Conf.	Pos.	Stability required $\geq 15^\circ$
4.2	5.4	Backward-direction stability		18,5 ⁰ Conf.	Pos.	Stability required $\geq 7^\circ$
4.2	5.5	Sideway-direction stability		4,5 ⁰ Conf.	Pos.	Stability required $\geq 3,5^\circ$
4.2	5.6	Stability – with loaded basket, bag, drip, oxygen cylinder	forwards	19,3 ⁰ Conf.	Pos.	Stability required $\geq 15^\circ$
			backwards	23,5 ⁰ Conf.	Pos.	Stability required $\geq 7^\circ$
			side	6,7 ⁰ Conf.	Pos.	Stability required $\geq 3,5^\circ$
4.3	V/I	Brakes	Servicing facility during rollator motion with more than 2 wheels	Conf.	Pos.	
	V/I		Parking brakes in rollator with more than 2 wheels and resting seat or intended for outdoor use	Conf.	Pos.	
	5.7.1.1		Brake grip distance (fig. 4, dimension 1)	65 mm Conf.	Pos.	≤ 75 mm
	5.7.1		Running brake effectiveness	Conf.	Pos.	Movement of rollator ≤ 10 mm in 1 minute
	Measur.		Force to set parking brake	30N Conf.	Pos.	≤ 60 N
	Measur.		Force to release parking brake	10N Conf.	Pos.	≤ 40 N
	5.7.2		Parking brake effectiveness	Conf.	Pos.	Movement of rollator ≤ 10 mm in 1 minute
	V/I		Possibility to compensate brake wear	Conf.	Pos.	
V/I	Brake not adversely affected by folding, unfolding or adjusting actions of rollator	Conf.	Pos.			
4.4	Measur. V/I	Handgrip		37 mm Conf.	Pos.	Width of handgrip ≥ 20 mm and ≤ 50 mm
4.5	Measur. V/I	Leg section and tip		--	N/A	ϕ tip ≥ 35 mm (tested rollator is equipped in four wheels)
4.6	5.10	Resting seat – static loading durability		Conf.	Pos.	1 minute under load 1,2 x user's weight $\pm 2\%$ (240kg)
4.7	5.12	Mechanical durability	Fatigue test	Conf.	Pos.	200 000 cycles with load. 160kg $\pm 2\%$, f=1Hz
4.7	5.11		Static loading test	Conf.	Pos.	loading 240kg $\pm 2\%$, 5sek.
4.8	V/I	Adjusting devices		Conf.	Pos.	
4.9	5.14	Folding mechanism		Conf.	Pos.	
4.11	ISO 10993-1	Materials and finish	Biocompatibility of material with human body	--	N/T	
	V/I		Free of discolouring of skin or clothing in contact with rollator materials	Conf.	Pos.	
	V/I		Burrs, shar edges, projections	Conf.	Pos.	
Marking and labelling of product						
6.2	V/I	a) Maximum user mass		--	N/T	
		b) Maximum safe working load (SWL) to be marked on accessories		--	N/T	
		c) Maximum allowed angle between the longitudinal centreline of the handle and the direction of motion, if the handles are sideways adjustable		--	N/T	
		d) Manufacturer's name or trade name and address		--	N/T	
		e) Manufacturer's model identification name and/or number		--	N/T	
		f) Month and year of manufacture		--	N/T	
		g) Maximum extension of the height adjustment, marked on the adjusting members		--	N/T	
		h) Maximum width of the rollator		--	N/T	
		i) Rollator intended for outdoor/indoor use		--	N/T	
4.10	V/I	Warning showing allowed angle between handle axis and direction of movement or physical stop of angle adjusting		--	N/T	

Contents of user manual and/or assembly manual or clear and indelible marking of product

6.3	V/I	a) Maximum rollator height	--	N/T
		b) Minimum rollator height	--	N/T
		c) maintenance and cleaning instructions, including a description of the method and suitable cleaning agents and any precautions needed to avoid corrosion and/or ageing of the materials used in construction of the rollator	--	N/T
		d) Instructions for assembly, adjustment of all kinds, folding and unfolding	--	N/T
		e) Warnings and advice about precautions relating to safe distances between moving and stationary parts (see EN 12182, Clauses 12 and 13, for guidance)	--	N/T
		f) Maximum safe working load (SWL) for load carrying accessories such as basket, tray, shopping bag, etc.	--	N/T
4.10	V/I	Warning in user manual on consequences of such an adjustment of angle between handle longitudinal axis and direction of movement outside allowed value (when handles are adjustable aside).	--	N/T

TEST CONDITIONS

Ambient temperature	19°C	Required temperature 21°C ±5°C
Relative humidity of air:	55 %	Not required
Comments:		
All tests performed with maximum height adjustment of rollator.		
All tests performed in the least stable position of self-adjusting wheels.		
Tests performed with handles positioned at their maximum (allowed) angle to the direction of motion (when adjustment is possible).		
Sequence of tests: stability test, static loading test, fatigue test.		
One rollator was tested.		
During visual inspection before testing any visible defects that could have influence on test results were not stated.		

Pos. – positive; Neg – negative; N/T – not tested; N/A – not applicable; N/R – not required, N/O – not occurred, V/I.- visual inspection, Conf.- conformed.

NOTE 1: Deformation – 35 mm, elastic deformation – 34 mm, permanent deformation – 1 mm.

CONCLUSIONS:

Testing object **conforms** with requirements of PN-EN ISO 11199-2 : 2005, excluding clause 4.10, 6.2, 6.3 in scope of mechanical testing ordered by client, excluding biocompatibility tests of material with human body according to PN-EN ISO 10993-1:2010.

---- END ----