

## EU Declaration of Conformity



According to Annex IV of the Medical Device Regulation 2017/745 concerning medical devices  
**IndeeLift** with head office at **C/Londres, 54, Mezz. 1 B; 08036 Barcelona** phone **+34 690 151 187**  
and email [international@indeelift.com](mailto:international@indeelift.com)

**SRN code:** ES-MF-000018968

**Declare under our sole responsibility that the system distributed under the brand name**

**PPU-S**

**UDI-DI basic:** 8437024311PPU-S58

Meet the provisions of the Medical Device Regulation 2017/745 concerning medical devices,  
which apply to it: the products are intended to be used as human floor lift.

Therefore, are class I medical device according to rule #1 of Annex VIII of the Medical Device  
Regulation 2017/745.

**Other standards applied:**

EN ISO 14971:2019

UNE-EN 60601-1:2006+A1:2013

UNE-EN 60601-1-2:2015+A1:2020

UNE-EN 60601-1-11:2015+A1:2021

**Name and function of the authorized person:**

Steven Jaw Powell. CEO

**Signature**

**Place and date:** Barcelona, 4<sup>th</sup> of April 2023