

Test report PB - 21 - 039 - AR - PA137 - 2 - E

Assistive products for persons with disability
Page 1/31, 2021-08-25



Test institute: Berlin Cert - Prüf- und Zertifizierstelle für Medizinprodukte GmbH,
Dovestraße 6, 10587 Berlin, Germany
Tel.: +49(0)30-314-25111, Fax: -23719

Client: Permobil Seating & Positioning
1501 S. 74th Street, IL 62223 Belleville, United States

Manufacturer: Permobil Seating & Positioning
1501 S. 74th Street, IL 62223 Belleville, United States

Test sample: Product type: Seat cushion
Model designation: ROHO Hybrid Select
Identification: HS1818c
Date of receipt / delivery condition: 2021-03-22 / new



Test standard: DIN EN 12182:2012
Assistive products for persons with disability – General requirements and test methods

Accreditation: Accredited by "Deutsche Akkreditierungsstelle GmbH" (DAkkS)
Reg.no.: D-PL-17591-01-00

Scope of testing: Complete test

Test period: 2021-03-31 to 2021-08-20

Test location: Rooms of the test institute

Test result: **The test requirements are fulfilled.**

Established: 2021-08-25

Released: 2021-08-25


M. Reiff
Test Technician


Dipl.-Ing. T. Hansen
Head of Rehabilitation Test Dept.

This test report may only be quoted in full. Any use for advertising purpose must be granted in writing.
This report is the result of a single examination of the object in question and is not generally applicable evaluation of the other products in regular production.

FB151_20170221R1.1

