

# Test report PB - 21 - 040 - R - PA023 - 2 - E

Hoists for the transfer of disabled persons


Page 1/21, 2021-11-05

**Test institute:** Berlin Cert - Prüf- und Zertifizierstelle für Medizinprodukte GmbH,  
Dovestr. 6, D-10587 Berlin  
Tel.: +49(0)30-314-25111, Fax: -23719

**Client:** VELA Vermund Larsen,  
Gøteborgvej 8-12, 9200 Aalborg, Danmark

**Manufacturer:** VELA Vermund Larsen,  
Gøteborgvej 8-12, 9200 Aalborg, Danmark

**Test sample:** Product type Car / wall mounted hoist  
Model designation: VELA Person Lift  
Identification: (01) 057 14339056847 (21) 410729  
Receipt / condition: 2021-04-20 / new



**Test standard:** DIN EN ISO 10535:2007  
Hoists for the transfer of disabled persons - Requirements and test methods

**Annotation:** The numbering corresponds to the applied standard  
DIN EN ISO 10535:2007

**Scope of testing:** Complete test

**Test period:** 2021-04-20 to 2021-10-20

**Test location:** Rooms of the test institute

**Environmental conditions:** Temperature =  $(21 \pm 5)$  °C; humidity =  $(50 \pm 20)$  %rH


**Test result:** **The test requirements are fulfilled.**

established: 2021-11-05

released: 2021-11-05



P. Rusch  
Test engineer



Dipl.-Ing. T. Hansen  
Test engineer

This test report may only be quoted in full. Any use for advertising purpose must be granted in writing.  
This report is the result of a single examination of the object in question and is not generally applicable evaluation of the other products in regular production.