


## EU Declaration of Conformity

**Manufacturer:**

iTs Designs Ltd.  
(T/A: Alert-iT Care Alarms)  
Fernie House (Unit 3 Ground Floor)  
Coalville Business Park  
Coalville, Leicestershire  
LE67 3NR, England  
Tel: +44 (0) 1530 239 900

<b>Description:</b>	The <b>"Plesio Pager"</b> is a basic pager for one-to-one care. It utilises Alert-iT's failsafe "Safelink™" radio protocol and can be paired with up to 3 devices. It will alert the caregiver to alarms, faults and failures of the radio network.	
<b>Model #:</b>	P168	
<b>Accessories:</b>	P153x P146	9v Power Supply for Pager USB Docking Station
<b>Directives:</b>	2014/53/EU 2014/30/EU 2011/65/EU + 2015/863 2012/19/EU	Radio Equipment Directive (RED) Electromagnetic Compatibility Directive (EMCD) Restriction of Hazardous Substances Directive (RoHS2) Waste Electrical and Electronic Equipment Directive (WEEE)
<b>Standards Applied:</b>	EN 60601-1:2006+A12:2014 EN 60601-1-2:2015 EN 55011:2016 EN 300 220-1 v3.1.1 EN 300 220-2 v3.1.1	General Requirements for Safety General Requirements for EMC EMC Standard - ISM Limits Radio Requirements - Short Range Devices Radio Requirements (434 MHz Model)

We hereby, under our sole responsibility declare, that the products listed above, are in conformity with the Directives and the relevant European harmonised standards listed above.



Date: 01/06/2021

Place: Leicester, UK.

Name: Tony Bullars

Position: Director