



TR9650

Mobile bath lift system

*Hygiene Equipment
Designed for People*



Declaration of Conformity

Valid from:	2021-04-01
Manufacturer:	TR Equipment AB Rundelgatan 2 573 35 Tranås Sweden
SRN:	SE-MF-000027486
Notified body:	N/A
Statement:	This declaration of conformity is issued under the sole responsibility of TR Equipment AB
Basic UDI-DI:	7350016060406J
Product:	TR 9650 – Mobile patient bath lift system
Product variants:	TR9650 Bath Chair lift, TR9650 Bath stretcher lift, TR9650 Bath Combi lift with chair, Stretcher side parts for combi lift with chair.
Presumption of compliance	ISO 10535:2006
Classification:	MDR Class I
MDR Annex:	VIII

We hereby declare that above mentioned devices comply with the European Medical Device Regulations MDR 2017/745, Swedish law 1993:584 regarding Medical devices, Regulation 1993:876 regarding Medical devices and regulation LVFS 2003:11 from the Swedish Medical Products Agency.

Tranås 2024-01-11

TR Equipment AB

Mico Olsen, CEO

Head office

TR Equipment AB, Sweden
Phone: +46 140 38 50 60
info@trequipment.se
www.trequipment.se

Group Subsidiaries

TR Group Inc, US, Phone: +1 (800) 752 6900
TRanspoRehaMed GmbH, Germany, Phone: +49 (0) 212 645850-0



TR EQUIPMENT