



Ref. Certif. No.

DE 3 - 40871

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product	General Medical Devices Sanitary aids
Name and address of the applicant	ArjoHuntleigh Polska Sp. z o.o. ul. Ks. Piotra Wawrzyniaka 2 62-052 Komorniki, POLAND
Name and address of the manufacturer	ArjoHuntleigh AB Hans Michelsensgatan 10, 211 20 Malmö, SWEDEN
Name and address of the factory	ArjoHuntleigh Polska Sp. z o.o. ul. Ks. Piotra Wawrzyniaka 2, 62-052 Komorniki, POLAND
Ratings and principal characteristics	Rated supply voltage: 12 Vdc (internally powered) Rated input power: max. 200 VA Type of applied part: BF Protection class: Internally powered Degree of protection: IP24 Max. user weight: 150kg / 330lbs Max. total load: 175kg / 385lbs Operatin time (ED): 10% (1 min. ON / 9 min. OFF)
Trade mark (if any)	ARJOHUNTLEIGH
Model/type Ref.	Concerto
Additional information (if necessary)	Certificate DE 3 - 40038 issued 2014-02-04 is replaced by this version due to non-technical changes
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012 IEC 60601-1-6:2010 IEC 60601-1-6:2010/AMD1:2013 IEC 60601-1-11:2015
as shown in the Test Report Ref. No. which forms part of this certificate	713120343

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date, 2017-12-21
CB 17 12 65428 017

Christoph Rappl



TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

Product Service