
Danish Centre

for technical aids for rehabilitation and education Gregersensvej - DK - 2630 Taastrup



TESTING REPORT

Technical aids for disabled persons - general requirements and test methods. according to EN 12182:1998

Walking aids manipulated by both arms - Requirements and test methods. according to prEN ISO/FDIS 11199-2:1997 (E)

Kravspesifikasjon ISO 12 18 12 Sparksykler
(None Powered Scooters, propelled by pushing with a foot)

Testing No. **98.032**

This page is page 1 of a total of 2 pages

This report should under no circumstances be reproduced, except in its total, without the permission in writing from the testing laboratory.

This report is not an approval of the product type tested, but a record of testing results for a specific product sample

Product tested : Petra, walking bike with tree wheels

Serial no.

Product received : 27.11.98

Testing notified by : Connie Hansen

Person to contact : Connie Hansen
Product delivered by : Connie Hansen

12-5-99 Date

Finn Såbye
Testing executive

Dept. of mobility, quality assurance and testing

Gregersensvej
DK- 2630 Taastrup
Tel : +45 43 99 33 22
Fax : +45 43 52 70 72
e-mail : daniscen@inet.uni-c.dk

SUMMARY OF THE TESTING

A sample walking bike of type Petra has been tested in accordance with all relevant requirements of the following sections of the EN 12182:1998 standard:

With the exception of:

- 5. Materials
- 5.1 Flammability
- 5.1.1 Upholstered parts
- 5.1.2 Upholstered parts
- 5.2 Biocompatibility and toxicity
- 5.3 Contaminants and residues

Which has not been requested.

The following requirements from the prEN ISO/FDIS 11199-2:1997 (E) was found to be relevant and has been tested.

- 5.3 Fatigue test
- 5.4 Static loading test
- 5.5 Forwards stability test
- 5.6 Backwards stability test
- 5.7 Sideways stability test

The following requirements from the ISO 12 18 12 was found to be relevant and has been tested.

- 3.4 Brakes
- 3 5 Handles

The testing took place during the period from 27.11.1998 to 11.05.1999 and was carried out under normal indoor conditions.

If not indicated in the testing protocol, measurements were made with tolerances which can be informed from the laboratory if desired.

TESTING RESULTS

All relevant requirements of the standards were found to be fulfilled.

The detailed testing results are recorded in the testing protocol attached. The parentheses refers to the remarks in the column "Requirements fulfilled".

Items marked with in the protocol includes results that are not a part of the accredited testing. It is estimated in these cases, however, that the bed/lift present no hazards when used in accordance with the manufactures instructions.

The result in those cases are therefore based on an estimation in relation to the intended use of the product, as specified in the instructions for use.