

CERTIFICATE

No. B 06 01 38696 001

Holder of Certificate: ATEC

Ingenieurbüro AG

Grepperstraße 77 6403 Küssnacht a. R. SWITZERLAND

Production Facility(ies):

38696

Certification Mark:



Product:

Components for wheelchairs

Model(s):

Wheelchair pulling unit Type: SWISS-TRAC SWT-1 / SWISS-TRAC KID SWT-2

Parameters:

max. obstacle height: max. safe slope:

120 mm 11°

max. speed:

forwards: 6 km/h, backwards 3,6 km/h

electro-motor: power electronics: Micro-Motor AG, type: PMS 110/40 ATEC Ingenieurbüro AG, type: SWT1

battery charger:

Power Charge A/S, model/type: GBCO7, 24 VDC, 8 A

max. trailing load:

130 kg (SWT-1)

batteries:

85 kg (SWT-2)

SWT-1:

MK Batteries, type: M40-122DG 2x (12 V, 40 Ah (20hr))

SWT-2:

Kung Long Batteries Industrial

Ltd., type: LG 28-12, 2x (12 V, 28 Ah (20hr))

Tested according to:

EN 12184:1999

(in Anlehnung / with regard to)

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. See also notes overleaf.

Test report no.:

70095172

Date. 2006-02-03

Page 1 of 1

