

## 11.2 OVERENSSTEMMELSESERKLÆRING Liftkar PT

SANO Transportgeraete GmbH påtager sig eneansvaret for at erklære,  
at følgende LIFTKAR-trappeklatre modeller er i overensstemmelse med:

EU-direktivet for medicinsk udstyr, klasse I EU2017/745 (MDR)

overensstemmelse.

Proceduren for overensstemmelsesvurdering i henhold til bilag IX  
(MDR).



LIFTKAR PT S 130/160

LIFTKAR PT Outdoor 120/150

LIFTKAR PT Universal 130/160

LIFTKAR PT Fold 130/160

LIFTKAR PT Plus 125

LIFTKAR Adapt 130/160

Eventuelle ændringer af produktet uden vores forudgående samtykke gør denne erklæring ugyldig.

Relevante standarder:

**ISO 7176-23:2002**

**ISO 7176-28:2012**

**ISO 7176-14:2008**

**ISO 7176-21:2009**

Requirements and test methods for attendant-operated stair-climbing devices

Wheelchairs - Part 28: Requirements and test methods for stair-climbing devices

Wheelchairs - Part 14: Power and control systems for electrically

powered wheelchairs scooters — Requirements and test methods

Requirements and test methods for electromagnetic compatibility of

electrically powered wheelchairs and scooters, and battery chargers

Denne overensstemmelseserklæring er gyldig, indtil en af de anførte typer ændres.



Ing. Jochum Bierma, Direktør