

Tests performed:

Certificate No. : C89202791
Report No. : 89202791-1
Project number 89202791

Properties and test methods	Requirement	Results
EN 12184: 2009	Type test	Pass

Full details of the results can be found in the following report(s):

Reference	Date of issue
TRN-Report 89202791-1	December 18, 2012

H. Fokkenrood

H. Ditmer

Author

Managing Director

hm.

CERTIFICATE

Certificate No. : C89202791
Report No. : 89202791-1
Project number 89202791

Holder: Medema Production A/S
Enggårdsvej 7, Snejbjerg
7400 Herning
Denmark

Product : Electrically powered wheelchair MC1122

Identification : MT12.36726.01

Date of test : Week 49 up to week 51, 2012


On basis of the results of the non-recurrent tests specified below, we declare that the above specimen fulfils the requirements of EN 12184:2009.

NB: This certificate itself shall not be considered as a type examination certificate.

H. Fokkenrood
Author



R. Ditmer
Managing Director



Page 1 of 2