



CENTRE FOR TESTING AND CERTIFICATION - MECH-TEST

Mechanical Laboratory

05-077 Warszawa-Wesoła, ul. Klonowa 22
tel.: +48 603 23-26-45, e-mail: cbc.mech.test@gmail.com, www.cbc.org.pl

Date 5.01.2016

TEST REPORT NO. **CBC-152/2015**

Page 1 of 4

Subject of testing:	<i>Walking aids with built-in handgrips and three or more legs of which two or more are having wheels, which provide support whilst walking</i>	Classification according to PN-EN ISO 9999:2011 : 12 06 06
Type / Model:	<i>Kudu Rollator MEDIUM</i>	Nr kat.: 312076 SN: G-563281
Manufacturer:	<i>MOBILEX A/S Nørskovvej 1 DK-8660 Skanderborg</i>	Number of specimens: 1
Applicant:	<i>A-Net s.c. 93-469 Łódź, ul. Łaskowice174</i>	
Kind of testing	<i>Mechanical testing for conformity with PN-EN ISO 11199-2 : 2005 excluding clause 4.10, 6.2, 6.3</i>	
Test started:	17.12.2015	
Test finished:	5.01.2016	

Approved by:

DYREKTOR

mgr inż. Andrzej Tkaczyk

Special comments / enclosures:

Copyright © 2012 by Centre for Testing and Certification (applicable to report form)

Test results refer only to tested units.

Test results reported here are not applicable to the further modifications of the product affecting its structure, material or technology.

This test report shall be neither copied differently as in the whole nor be published without written consent of the Laboratory.



CHARACTERISTIC OF PRODUCT

Name : <i>Kudu - MEDIUM</i>	Dimension of rollator: --
Nr. kat: <i>312076</i>	Product code: --
Maximum permissible user mass: <i>150 kg</i>	Mass of rollator: <i>7,905 kg</i>

Description		Comments	
Elements/parameters/materials/dimensions			
Dimensions of walking rollator (fig. 2 PN-EN ISO 11199-2)	Distance between handgrips (dimension 2)	440 mm	
	Angle between of handgrip axis and direction of movement (α)	0°	
	Height of rollator (dimension 6)	788 mm	min.
		881 mm	max.
	Width of rollator (dimension 5)	612 mm	
	Turning width (dimension 1)	832 mm	
Length of rollator (dimension 4)	685 mm		
Dimensions of folded rollator (mm)		679 x 788 x 271	
Fig. 3	Handgrip - diameter	33 mm	Anatomical handgrip
	Handgrip - length	100 mm	
Wheels of rollator	Front wheels- quantity	2	castor wheels
	Front wheels - diameter	201 mm	
	Front wheels - width	34 mm	
	Front wheels - brake	none	
	Rear wheels - quantity	2	
	Rear wheels - diameter	201 mm	
	Rear wheels - width	34 mm	
	Rear wheels - brake	Included	
Tip	Diameter		
	Material	Not any	
	Colour		
Material of rollator (fig. 1)	Front legs	Aluminum,	
	Bracing member (no. 8)	steel, plastic,	
	Rear legs	Bolts, nuts	
	Height adjusting device (no. 4)		
	Handgrip (no 5), Brake elements	plastic	

PHOTO OF PRODUCT




Mobilex **Kudu rollator**
 For outdoor and indoor use

MOBILEX A/S
 Grønlandsvej 5
 DK - 8660 Skanderborg
 Nørskovvej 1
 Tel: +45 87 93 22 20
 www.mobilex.dk

Max. 150 kg
 Max. width = 63 cm
 Max. load in bag = 5 kg

CE Produced Aug. 2014



RESULT OF TESTS ACCORDING TO PN-EN ISO 11199-2:2005

Requirements according to clause	Test method according to clause	Checked characteristics/assemblies/parameters	Real value	Test result	Comments	
4.1	Measur.	Manoeuvrability	\varnothing 201 mm width 34 mm Conf.	Pos.	\varnothing front wheels ≥ 75 mm <u>outdoor intended rollator:</u> \varnothing front wheels ≥ 180 mm width of wheels ≥ 22 mm	
4.2	5.3	Forward-direction stability	24,0 ⁰ Conf.	Pos.	Stability required $\geq 15^\circ$	
4.2	5.4	Backward-direction stability	12,3 ⁰ Conf.	Pos.	Stability required $\geq 7^\circ$	
4.2	5.5	Sideway-direction stability	5,2 ⁰ Conf.	Pos.	Stability required $\geq 3,5^\circ$	
4.2	5.6	Stability – with loaded bag /+tray	forwards	22,0 ⁰ /21,0 ⁰ Conf.	Pos.	Stability required $\geq 15^\circ$
			backwards	17,5 ⁰ /20,5 ⁰ Conf.	Pos.	Stability required $\geq 7^\circ$
			side	6,4 ⁰ /8,3 ⁰ Conf.	Pos.	Stability required $\geq 3,5^\circ$
4.3	V/I	Brakes	Servicing facility during rollator motion with more than 2 wheels	Conf.	Pos.	
	V/I		Parking brakes in rollator with more than 2 wheels and resting seat or intended for outdoor use	Conf.	Pos.	
	5.7.1.1		Brake grip distance (fig. 4, dimension 1)	75 mm Conf.	Pos.	≤ 75 mm
	5.7.1		Running brake effectiveness	Conf.	Pos.	Movement of rollator ≤ 10 mm in 1 minute
	Measur.		Force to set parking brake	30N Conf.	Pos.	≤ 60 N
	Measur.		Force to release parking brake	20N Conf.	Pos.	≤ 40 N
	5.7.2		Parking brake effectiveness	Conf.	Pos.	Movement of rollator ≤ 10 mm in 1 minute
	V/I		Possibility to compensate brake wear	Conf.	Pos.	
V/I	Brake not adversely affected by folding, unfolding or adjusting actions of rollator	Conf.	Pos.			
4.4	Measur. V/I	Handgrip	33 mm Conf.	Pos.	Width of handgrip ≥ 20 mm and ≤ 50 mm	
4.5	Measur. V/I	Leg section and tip	--	N/A	\varnothing tip ≥ 35 mm (tested rollator is equipped in four wheels)	
4.6	5.10	Resting seat – static loading durability	Conf.	Pos.	1 minute under load 1,2 x user's weight $\pm 2\%$ (180kg)	
4.7	5.12	Mechanical durability	Fatigue test	Conf.	Pos.	200 000 cycles with load. 120kg $\pm 2\%$, f=1Hz
4.7	5.11		Static loading test	Conf.	Pos.	loading 180kg $\pm 2\%$, 5sek.
4.8	V/I	Adjusting devices	Conf.	Pos.		
4.9	5.14	Folding mechanism	Conf.	Pos.		
4.11	ISO 10993-1	Materials and finish	Biocompatibility of material with human body	--	N/T	
	V/I		Free of discolouring of skin or clothing in contact with rollator materials	Conf.	Pos.	
	V/I		Burrs, shar edges, projections	Conf.	Pos.	
Marking and labelling of product						
6.2	V/I	a) Maximum user mass	--	N/T		
		b) Maximum safe working load (SWL) to be marked on accessories	--	N/T		
		c) Maximum allowed angle between the longitudinal centreline of the handle and the direction of motion, if the handles are sideways adjustable	--	N/T		
		d) Manufacturer's name or trade name and address	--	N/T		
		e) Manufacturer's model identification name and/or number	--	N/T		
		f) Month and year of manufacture	--	N/T		
		g) Maximum extension of the height adjustment, marked on the adjusting members	--	N/T		
		h) Maximum width of the rollator	--	N/T		
		i) Rollator intended for outdoor/indoor use	--	N/T		
4.10	V/I	Warning showing allowed angle between handle axis and direction of movement or physical stop of angle adjusting	--	N/T		

Contents of user manual and/or assembly manual or clear and indelible marking of product

6.3	V/I	a) Maximum rollator height	--	N/T
		b) Minimum rollator height	--	N/T
		c) maintenance and cleaning instructions, including a description of the method and suitable cleaning agents and any precautions needed to avoid corrosion and/or ageing of the materials used in construction of the rollator	--	N/T
		d) Instructions for assembly, adjustment of all kinds, folding and unfolding	--	N/T
		e) Warnings and advice about precautions relating to safe distances between moving and stationary parts (see EN 12182, Clauses 12 and 13, for guidance)	--	N/T
		f) Maximum safe working load (SWL) for load carrying accessories such as basket, tray, shopping bag, etc.	--	N/T
4.10	V/I	Warning in user manual on consequences of such an adjustment of angle between handle longitudinal axis and direction of movement outside allowed value (when handles are adjustable aside).	--	N/T

TEST CONDITIONS

Ambient temperature	19°C	Required temperature 21°C ±5°C
Relative humidity of air:	55 %	Not required
<i>Comments:</i>		
All tests performed with maximum height adjustment of rollator.		
All tests performed in the least stable position of self-adjusting wheels.		
Tests performed with handles positioned at their maximum (allowed) angle to the direction of motion (when adjustment is possible).		
Sequence of tests: stability test, static loading test, fatigue test.		
One rollator was tested.		
During visual inspection before testing any visible defects that could have influence on test results were not stated.		

Pos. – positive; Neg – negative; N/T – not tested; N/A – not applicable; N/R – not required, N/O – not occurred, V/I.- visual inspection, Conf.- conformed.

NOTE 1: Deformation – 50 mm, elastic deformation – 47 mm, permanent deformation – 3 mm.

NOTE 2: An additional fatigue tests on the two-drum stand with obstacles of a height of 12mm, according to ISO 7176-8: 2002, cl. 10.4.2., was performed.

- ▶ The handles were loaded with a load of 120 kg. After 15 000 cycles, the rollator was not damaged. The test result was positive.
- ▶ The seat was loaded with a load of 100 kg. After a further 35 000 cycles, the rollator was not damaged. The test result positive

NOTE 3: Additional Drop Pouch tests were performed. The seat was loaded with a mass of 100 kg. Rear of the rollator was dropped (on wheels) 5 times from a height of 100mm. Rollator was not damaged. The test result positive.

CONCLUSIONS:

Testing object **conforms** with requirements of PN-EN ISO 11199-2 : 2005, excluding clause 4.10, 6.2, 6.3 in scope of mechanical testing ordered by client, excluding biocompatibility tests of material with human body according to PN-EN ISO 10993-1:2010 .

---- END ----